

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102925\
 Data File : VN088149.D
 Acq On : 29 Oct 2025 11:06
 Operator : JC\MD
 Sample : VN1029WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1029WBL01

Quant Time: Oct 30 04:02:25 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102325W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 25 00:15:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	220088	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.083	114	494622	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	463889	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	222427	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.559	65	207466	54.513	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	109.020%
35) Dibromofluoromethane	8.153	113	154801	46.586	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.180%
50) Toluene-d8	10.547	98	569614	44.488	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	88.980%
62) 4-Bromofluorobenzene	12.829	95	217445	48.088	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	96.180%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102925\
Data File : VN088149.D
Acq On : 29 Oct 2025 11:06
Operator : JC\MD
Sample : VN1029WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1029WBL01

Quant Time: Oct 30 04:02:25 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102325W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 25 00:15:22 2025
Response via : Initial Calibration

