

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102925\
 Data File : VN088156.D
 Acq On : 29 Oct 2025 13:44
 Operator : JC\MD
 Sample : PB170296TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB170296TB

Quant Time: Oct 30 04:05:58 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102325W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 25 00:15:22 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	257074	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	592700	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.841	117	583875	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	280235	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.559	65	271649	61.109	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 122.220%		
35) Dibromofluoromethane	8.147	113	192003	48.220	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 96.440%		
50) Toluene-d8	10.547	98	710422	46.304	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 92.600%		
62) 4-Bromofluorobenzene	12.829	95	277866	51.282	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 102.560%		
Target Compounds						
					Qvalue	
16) Acetone	4.418	43	45691	20.459	ug/l	94
18) Methyl Acetate	5.024	43	10457	2.287	ug/l	99
20) Methylene Chloride	5.265	84	14096	4.765	ug/l #	87
43) Isopropyl Acetate	8.671	43	28258	2.451	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102925\
Data File : VN088156.D
Acq On : 29 Oct 2025 13:44
Operator : JC\MD
Sample : PB170296TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB170296TB

Quant Time: Oct 30 04:05:58 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102325W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 25 00:15:22 2025
Response via : Initial Calibration

