


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN103018\
Data File : VN052054.D
Acq On : 30 Oct 2018 7:41
Operator : MD\SY
Sample : BFB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 1 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N103018W.M
Title : SW846 8260
Last Update : Wed Oct 31 02:45:22 2018



AutoFind: Scans 3304, 3305, 3306; Background Corrected with Scan 3289

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.5	12176	PASS
75	95	30	60	54.5	30872	PASS
95	95	100	100	100.0	56642	PASS
96	95	5	9	6.7	3782	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	63.7	36061	PASS
175	174	5	9	6.8	2437	PASS
176	174	95	101	95.1	34277	PASS
177	176	5	9	8.0	2735	PASS