

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN103019\
 Data File : VN059018.D
 Acq On : 30 Oct 2019 10:32
 Operator : JC/SP
 Sample : PB124402TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB124402TB

Quant Time: Oct 31 04:45:25 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 19 00:16:20 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	413702	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	659527	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	604803	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	216046	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	283974	46.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.60%	
35) Dibromofluoromethane	7.57	113	213869	49.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.64%	
50) Toluene-d8	10.08	98	832844	49.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.96%	
62) 4-Bromofluorobenzene	12.40	95	267899	43.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.68%	
Target Compounds						
16) Acetone	3.80	43	15188	5.066	ug/l	94
20) Methylene Chloride	4.53	84	17465	3.411	ug/l	89
43) Isopropyl Acetate	8.15	43	123887	10.499	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN103019\
Data File : VN059018.D
Acq On : 30 Oct 2019 10:32
Operator : JC/SP
Sample : PB124402TB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB124402TB

Quant Time: Oct 31 04:45:25 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 19 00:16:20 2019
Response via : Initial Calibration

