

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN103019\
 Data File : VN059020.D
 Acq On : 30 Oct 2019 11:16
 Operator : JC/SP
 Sample : K5148-17
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OR-03-102919

Quant Time: Oct 31 04:58:07 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 19 00:16:20 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	374819	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	579937	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	532152	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	197979	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	253244	46.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.14%	
35) Dibromofluoromethane	7.57	113	190923	50.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.14%	
50) Toluene-d8	10.08	98	740086	50.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.00%	
62) 4-Bromofluorobenzene	12.40	95	237974	43.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.56%	
Target Compounds						
16) Acetone	3.80	43	14737	5.425	ug/l	95
18) Methyl Acetate	4.31	43	6305	1.021	ug/l	95
20) Methylene Chloride	4.53	84	286067	61.674	ug/l	96
43) Isopropyl Acetate	8.15	43	100078	9.646	ug/l #	91
95) Naphthalene	15.13	128	17941	3.737	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN103019\
Data File : VN059020.D
Acq On : 30 Oct 2019 11:16
Operator : JC/SP
Sample : K5148-17
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OR-03-102919

Quant Time: Oct 31 04:58:07 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 19 00:16:20 2019
Response via : Initial Calibration

