

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN103019\
 Data File : VN059022.D
 Acq On : 30 Oct 2019 12:30
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 31 05:05:47 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 19 00:16:20 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	96	0.00
2 T	Dichlorodifluoromethane	0.623	0.532	14.6	83	0.00
3 P	Chloromethane	0.772	0.682	11.7	87	0.00
4 C	Vinyl Chloride	0.767	0.704	8.2#	89	0.00
5 T	Bromomethane	0.480	0.408	15.0	82	0.00
6 T	Chloroethane	0.492	0.436	11.4	88	0.00
7 T	Trichlorofluoromethane	0.949	0.885	6.7	89	0.00
8 T	Diethyl Ether	0.328	0.343	-4.6	104	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.526	0.518	1.5	99	0.00
10 T	Methyl Iodide	0.692	0.639	7.7	89	0.00
11 T	Tert butyl alcohol	0.122	0.121	0.8	99	0.00
12 CM	1,1-Dichloroethene	0.493	0.481	2.4#	95	0.00
13 T	Acrolein	0.084	0.072	14.3	80	0.00
14 T	Allyl chloride	0.997	0.924	7.3	89	0.00
15 T	Acrylonitrile	0.306	0.328	-7.2	100	0.00
16 T	Acetone	0.362	0.370	-2.2	94	0.00
17 T	Carbon Disulfide	1.588	1.354	14.7	83	0.00
18 T	Methyl Acetate	0.824	0.881	-6.9	109	0.00
19 T	Methyl tert-butyl Ether	1.726	1.856	-7.5	101	0.00
20 T	Methylene Chloride	0.619	0.601	2.9	101	0.00
21 T	trans-1,2-Dichloroethene	0.544	0.530	2.6	94	0.00
22 T	Diisopropyl ether	1.951	2.115	-8.4	101	0.00
23 T	Vinyl Acetate	1.567	1.739	-11.0	96	0.00
24 P	1,1-Dichloroethane	1.131	1.126	0.4	98	0.00
25 T	2-Butanone	0.470	0.485	-3.2	95	0.00
26 T	2,2-Dichloropropane	0.978	0.991	-1.3	99	0.00
27 T	cis-1,2-Dichloroethene	0.622	0.633	-1.8	97	0.00
28 T	Bromochloromethane	0.377	0.377	0.0	97	0.00
29 T	Tetrahydrofuran	0.288	0.302	-4.9	96	0.00
30 C	Chloroform	1.135	1.106	2.6#	97	0.00
31 T	Cyclohexane	1.054	0.963	8.6	90	0.00
32 T	1,1,1-Trichloroethane	0.956	0.936	2.1	94	0.00
33 S	1,2-Dichloroethane-d4	0.733	0.690	5.9	91	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	95	0.00
35 S	Dibromofluoromethane	0.329	0.337	-2.4	97	0.00
36 T	1,1-Dichloropropene	0.512	0.527	-2.9	94	0.00
37 T	Ethyl Acetate	0.558	0.586	-5.0	94	0.00
38 T	Carbon Tetrachloride	0.527	0.512	2.8	92	0.00
39 T	Methylcyclohexane	0.575	0.589	-2.4	92	0.00
40 TM	Benzene	1.503	1.572	-4.6	97	0.00
41 T	Methacrylonitrile	0.248	0.261	-5.2	102	0.00
42 TM	1,2-Dichloroethane	0.579	0.571	1.4	93	0.00
43 T	Isopropyl Acetate	0.895	0.951	-6.3	97	0.00
44 TM	Trichloroethene	0.366	0.374	-2.2	96	0.00
45 C	1,2-Dichloropropane	0.412	0.437	-6.1#	99	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN103019\
 Data File : VN059022.D
 Acq On : 30 Oct 2019 12:30
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 31 05:05:47 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 19 00:16:20 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.261	0.265	-1.5	96	0.00
47 T	Bromodichloromethane	0.537	0.559	-4.1	97	0.00
48 T	Methyl methacrylate	0.406	0.439	-8.1	95	0.00
49 T	1,4-Dioxane	0.007	0.008	-14.3	106	0.00
50 S	Toluene-d8	1.276	1.267	0.7	93	0.00
51 T	4-Methyl-2-Pentanone	0.522	0.597	-14.4	97	0.00
52 CM	Toluene	0.891	0.956	-7.3#	96	0.00
53 T	t-1,3-Dichloropropene	0.595	0.641	-7.7	97	0.00
54 T	cis-1,3-Dichloropropene	0.635	0.688	-8.3	97	0.00
55 T	1,1,2-Trichloroethane	0.367	0.387	-5.4	100	0.00
56 T	Ethyl methacrylate	0.540	0.624	-15.6	100	0.00
57 T	1,3-Dichloropropane	0.645	0.683	-5.9	98	0.00
58 T	2-Chloroethyl Vinyl ether	0.234	0.264	-12.8	98	0.00
59 T	2-Hexanone	0.404	0.458	-13.4	95	0.00
60 T	Dibromochloromethane	0.395	0.434	-9.9	101	0.00
61 T	1,2-Dibromoethane	0.376	0.401	-6.6	98	0.00
62 S	4-Bromofluorobenzene	0.469	0.491	-4.7	98	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	94	0.00
64 T	Tetrachloroethene	0.350	0.353	-0.9	94	0.00
65 PM	Chlorobenzene	1.018	1.102	-8.3	99	0.00
66 T	1,1,1,2-Tetrachloroethane	0.384	0.420	-9.4	101	0.00
67 C	Ethyl Benzene	1.786	2.024	-13.3#	97	0.00
68 T	m/p-Xylenes	0.662	0.751	-13.4	97	0.00
69 T	o-Xylene	0.632	0.721	-14.1	99	0.00
70 T	Styrene	1.030	1.217	-18.2	100	0.00
71 P	Bromoform	0.298	0.336	-12.8	103	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	97	0.00
73 T	Isopropylbenzene	3.654	3.918	-7.2	99	0.00
74 T	N-amyl acetate	1.657	1.778	-7.3	94	0.00
75 P	1,1,2,2-Tetrachloroethane	1.285	1.311	-2.0	99	0.00
76 T	1,2,3-Trichloropropane	1.159	1.237	-6.7	107	0.00
77 T	Bromobenzene	0.901	0.940	-4.3	101	0.00
78 T	n-propylbenzene	4.319	4.707	-9.0	99	0.00
79 T	2-Chlorotoluene	2.644	2.762	-4.5	99	0.00
80 T	1,3,5-Trimethylbenzene	3.070	3.342	-8.9	97	0.00
81 T	trans-1,4-Dichloro-2-butene	0.445	0.472	-6.1	99	0.00
82 T	4-Chlorotoluene	2.702	2.849	-5.4	98	0.00
83 T	tert-Butylbenzene	2.625	2.848	-8.5	100	0.00
84 T	1,2,4-Trimethylbenzene	3.028	3.364	-11.1	97	0.00
85 T	sec-Butylbenzene	3.500	3.871	-10.6	99	0.00
86 T	p-Isopropyltoluene	3.088	3.508	-13.6	99	0.00
87 T	1,3-Dichlorobenzene	1.610	1.704	-5.8	100	0.00
88 T	1,4-Dichlorobenzene	1.612	1.675	-3.9	98	0.00
89 T	n-Butylbenzene	2.837	3.165	-11.6	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN103019\
Data File : VN059022.D
Acq On : 30 Oct 2019 12:30
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Oct 31 05:05:47 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 19 00:16:20 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.626	0.641	-2.4	99	0.00
91 T	1,2-Dichlorobenzene	1.583	1.695	-7.1	100	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.291	0.264	9.3	90	0.00
93 T	1,2,4-Trichlorobenzene	0.873	0.987	-13.1	99	0.00
94 T	Hexachlorobutadiene	0.513	0.554	-8.0	104	0.00
95 T	Naphthalene	2.436	2.643	-8.5	91	0.00
96 T	1,2,3-Trichlorobenzene	0.858	0.962	-12.1	100	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6