

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN103020\
 Data File : VN064389.D
 Acq On : 30 Oct 2020 10:30
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050

Quant Time: Oct 30 14:35:13 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N102720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 27 12:23:17 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	248671	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	396204	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	376925	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	174259	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	7.98	65	189579	47.44	ug/l	0.00
Spiked Amount			Recovery	=	94.88%	
35) Dibromofluoromethane	7.54	113	130202	49.28	ug/l	0.00
Spiked Amount			Recovery	=	98.56%	
50) Toluene-d8	10.05	98	477814	48.89	ug/l	0.00
Spiked Amount			Recovery	=	97.78%	
62) 4-Bromofluorobenzene	12.37	95	184112	48.32	ug/l	0.00
Spiked Amount			Recovery	=	96.64%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) Dichlorodifluoromethane	1.83	85	148704	52.022	ug/l	99
3) Chloromethane	2.04	50	150617	48.744	ug/l	99
4) Vinyl Chloride	2.16	62	157815	52.802	ug/l	97
5) Bromomethane	2.53	94	94744	49.058	ug/l	100
6) Chloroethane	2.67	64	93091	52.125	ug/l	100
7) Trichlorofluoromethane	2.98	101	283748	56.670	ug/l	99
8) Diethyl Ether	3.37	74	86810	54.806	ug/l	94
9) 1,1,2-Trichlorotrifluoroet	3.72	101	136431	54.213	ug/l	98
10) Methyl Iodide	3.91	142	154866	46.115	ug/l	99
11) Tert butyl alcohol	4.72	59	123915	213.610	ug/l	99
12) 1,1-Dichloroethene	3.69	96	124541	52.421	ug/l	94
13) Acrolein	3.57	56	24842	201.061	ug/l	96
14) Allyl chloride	4.27	41	232998	48.031	ug/l	94
15) Acrylonitrile	4.92	53	310960	249.269	ug/l	99
16) Acetone	3.77	43	405835	299.337	ug/l	98
17) Carbon Disulfide	4.01	76	347996	49.508	ug/l	99
18) Methyl Acetate	4.27	43	146451	48.787	ug/l	100
19) Methyl tert-butyl Ether	4.99	73	481516	52.750	ug/l	98
20) Methylene Chloride	4.50	84	145187	51.230	ug/l	98
21) trans-1,2-Dichloroethene	4.99	96	139778	52.741	ug/l	92
22) Diisopropyl ether	5.89	45	513674	54.174	ug/l	98
23) Vinyl Acetate	5.84	43	2179701	271.719	ug/l	100
24) 1,1-Dichloroethane	5.79	63	277921	52.381	ug/l	99
25) 2-Butanone	6.78	43	528715	263.507	ug/l	99
26) 2,2-Dichloropropane	6.77	77	275817	53.116	ug/l	100
27) cis-1,2-Dichloroethene	6.78	96	160705	52.984	ug/l	97
28) Bromochloromethane	7.14	49	125875	49.794	ug/l	99
29) Tetrahydrofuran	7.16	42	304353	233.655	ug/l	100
30) Chloroform	7.32	83	302209	53.362	ug/l	100
31) Cyclohexane	7.61	56	218914	47.201	ug/l	96
32) 1,1,1-Trichloroethane	7.52	97	285242	52.451	ug/l	99
36) 1,1-Dichloropropene	7.75	75	217937	54.503	ug/l	99
37) Ethyl Acetate	6.87	43	202240	50.072	ug/l	99
38) Carbon Tetrachloride	7.73	117	258263	55.711	ug/l	98

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN103020\
 Data File : VN064389.D
 Acq On : 30 Oct 2020 10:30
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050

Quant Time: Oct 30 14:35:13 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N102720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 27 12:23:17 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
39) Methylcyclohexane	9.04	83	186402	47.014	ug/l	98
40) Benzene	8.00	78	598429	53.766	ug/l	97
41) Methacrylonitrile	7.12	41	100924	50.337	ug/l	92
42) 1,2-Dichloroethane	8.08	62	267363	55.608	ug/l	99
43) Isopropyl Acetate	8.12	43	367481	53.225	ug/l #	91
44) Trichloroethene	8.80	130	161767	55.481	ug/l	92
45) 1,2-Dichloropropane	9.08	63	160077	53.708	ug/l	97
46) Dibromomethane	9.17	93	113282	54.329	ug/l	97
47) Bromodichloromethane	9.37	83	246469	53.337	ug/l	98
48) Methyl methacrylate	9.16	41	176738	49.615	ug/l	99
49) 1,4-Dioxane	9.16	88	49410	941.060	ug/l	96
51) 4-Methyl-2-Pentanone	9.95	43	1028266	256.209	ug/l	99
52) Toluene	10.12	92	384111	54.690	ug/l	100
53) t-1,3-Dichloropropene	10.35	75	274725	56.227	ug/l	100
54) cis-1,3-Dichloropropene	9.80	75	277796	55.233	ug/l	98
55) 1,1,2-Trichloroethane	10.53	97	151310	54.578	ug/l	99
56) Ethyl methacrylate	10.40	69	224449	53.482	ug/l	99
57) 1,3-Dichloropropane	10.68	76	262023	55.405	ug/l	99
58) 2-Chloroethyl Vinyl ether	9.66	63	579324	252.967	ug/l	99
59) 2-Hexanone	10.72	43	786313	266.019	ug/l	98
60) Dibromochloromethane	10.87	129	186104	56.775	ug/l	99
61) 1,2-Dibromoethane	10.97	107	159946	56.621	ug/l	98
64) Tetrachloroethene	10.60	164	154529	49.881	ug/l	98
65) Chlorobenzene	11.40	112	413076	54.057	ug/l	100
66) 1,1,1,2-Tetrachloroethane	11.48	131	164565	53.282	ug/l	97
67) Ethyl Benzene	11.48	91	753168	53.015	ug/l	96
68) m/p-Xylenes	11.59	106	557044	105.804	ug/l	100
69) o-Xylene	11.92	106	270002	53.231	ug/l	97
70) Styrene	11.94	104	457173	54.972	ug/l	99
71) Bromoform	12.10	173	124850	51.504	ug/l #	100
73) Isopropylbenzene	12.22	105	698539	53.522	ug/l	99
74) N-amyl acetate	12.04	43	318677	55.473	ug/l	99
75) 1,1,2,2-Tetrachloroethane	12.48	83	199575	52.607	ug/l	98
76) 1,2,3-Trichloropropane	12.52	75	294366	70.540	ug/l #	100
77) Bromobenzene	12.50	156	175103	54.789	ug/l	99
78) n-propylbenzene	12.56	91	791891	53.593	ug/l	99
79) 2-Chlorotoluene	12.65	91	500073	52.943	ug/l	100
80) 1,3,5-Trimethylbenzene	12.70	105	578757	52.407	ug/l	99
81) trans-1,4-Dichloro-2-buten	12.27	75	76343	52.697	ug/l	96
82) 4-Chlorotoluene	12.74	91	533399	55.433	ug/l	99
83) tert-Butylbenzene	12.96	119	459444	50.699	ug/l	99
84) 1,2,4-Trimethylbenzene	13.01	105	586330	53.135	ug/l	99
85) sec-Butylbenzene	13.14	105	566741	48.328	ug/l	100
86) p-Isopropyltoluene	13.26	119	524144	49.612	ug/l	99
87) 1,3-Dichlorobenzene	13.25	146	301633	53.912	ug/l	98
88) 1,4-Dichlorobenzene	13.33	146	302593	53.101	ug/l	99
89) n-Butylbenzene	13.59	91	444453	47.754	ug/l	97
90) Hexachloroethane	13.85	117	96875	48.825	ug/l	98
91) 1,2-Dichlorobenzene	13.62	146	296466	53.298	ug/l	98
92) 1,2-Dibromo-3-Chloropropan	14.24	75	47451	50.408	ug/l	99


Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN103020\
Data File : VN064389.D
Acq On : 30 Oct 2020 10:30
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Quant Time: Oct 30 14:35:13 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N102720W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 27 12:23:17 2020
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
93) 1,2,4-Trichlorobenzene	14.88	180	145051	45.559	ug/l	98
94) Hexachlorobutadiene	14.98	225	69903	40.952	ug/l	98
95) Naphthalene	15.10	128	425018	43.472	ug/l	99
96) 1,2,3-Trichlorobenzene	15.28	180	134673	43.978	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.62 min Scan# 1829

Delta R.T. -0.00 min

Lab File: VN064389.D

Acq: 30 Oct 2020 10:30

Instrument :

MSVOA_N

ClientSampleId :


VSTDCCC050

Tot Ion:168 Resp: 248671

Ion Ratio Lower Upper

168 100

99 67.9 52.7 79.1



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VNO

7.62

100000

80000



60000

40000

20000

0

Time-->



#2

Dichlorodifluoromethane

Concen: 52.022 ug/l

RT: 1.83 min Scan# 28

Delta R.T. -0.00 min

Lab File: VN064389.D


Acq: 30 Oct 2020 10:30

Tgt Ion: 85 Resp: 148704

Ion Ratio Lower Upper

85 100

87 32.6 16.0 47.9



Abundance Ion 84.90 (84.60 to 85.60): VNO

Ion 86.90 (86.60 to 87.60): VNO

1.83

120000

100000

80000


60000


40000

20000

0

Time-->





#3

Chloromethane

Concen: 48.744 ug/l

RT: 2.04 min Scan# 92

Delta R.T. -0.00 min

Lab File: VN064389.D

Acq: 30 Oct 2020 10:30

Instrument :

MSVOA_N

ClientSampleId :


VSTDCCC050

Tot Ion: 50 Resp: 150617




Ion Ratio Lower Upper

50 100

52 31.4 25.6 38.4



Abundance Ion 49.90 (49.60 to 50.60): VNO
Ion 51.90 (51.60 to 52.60): VNO



#4

Vinyl Chloride

Concen: 52.802 ug/l

RT: 2.16 min Scan# 132

Delta R.T. 0.00 min

Lab File: VN064389.D


Acq: 30 Oct 2020 10:30

Tgt Ion: 62 Resp: 157815



Ion Ratio Lower Upper


62 100

64 31.3 26.2 39.4



Abundance Ion 61.90 (61.60 to 62.60): VNO
Ion 63.90 (63.60 to 64.60): VNO





#5

Bromomethane

Concen: 49.058 ug/l

RT: 2.53 min Scan# 245

Delta R.T. 0.00 min

Lab File: VN064389.D

Acq: 30 Oct 2020 10:30

Instrument :

MSVOA_N

ClientSampleId :


VSTDCCC050

Tot Ion: 94 Resp: 94744

Ion Ratio Lower Upper




94 100

96 92.4 73.6 110.4



Abundance Ion 93.90 (93.60 to 94.60): VN0

Ion 95.90 (95.60 to 96.60): VN0



#6

Chloroethane

Concen: 52.125 ug/l

RT: 2.67 min Scan# 289

Delta R.T. 0.00 min

Lab File: VN064389.D


Acq: 30 Oct 2020 10:30

Tgt Ion: 64 Resp: 93091

Ion Ratio Lower Upper



64 100


66 32.9 26.5 39.7



Abundance Ion 63.90 (63.60 to 64.60): VN0

Ion 65.90 (65.60 to 66.60): VN0






#7




Trichlorofluoromethane
Concen: 56.670 ug/l
RT: 2.98 min Scan# 387
Delta R.T. 0.00 min
Lab File: VN064389.D
Acq: 30 Oct 2020 10:30

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tot Ion:101 Resp: 283748
Ion Ratio Lower Upper
101 100
103 64.1 51.7 77.5




Abundance Ion 100.90 (100.60 to 101.60): V
Ion 102.80 (102.50 to 103.50): V





#8


Diethyl Ether
Concen: 54.806 ug/l
RT: 3.37 min Scan# 507
Delta R.T. -0.00 min
Lab File: VN064389.D
Acq: 30 Oct 2020 10:30

Tgt Ion: 74 Resp: 86810
Ion Ratio Lower Upper
74 100
45 96.8 51.3 154.0



Abundance Ion 74.00 (73.70 to 74.70): VNO
Ion 45.00 (44.70 to 45.70): VNO





#9

1,1,2-Trichlorotrifluoroethane

Concen: 54.213 ug/l

RT: 3.72 min Scan# 615

Delta R.T. 0.00 min

Lab File: VN064389.D

Acq: 30 Oct 2020 10:30

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050


Tot Ion:101 Resp: 136431

Ion Ratio Lower Upper

101 100

85 47.8 38.4 57.6




151 74.2 62.0 93.0



Abundance Ion 100.90 (100.60 to 101.60): V

Ion 84.90 (84.60 to 85.60): VN

Ion 150.80 (150.50 to 151.50): V



#10

Methyl Iodide

Concen: 46.115 ug/l

RT: 3.91 min Scan# 675

Delta R.T. -0.00 min

Lab File: VN064389.D

Acq: 30 Oct 2020 10:30


Tgt Ion:142 Resp: 154866

Ion Ratio Lower Upper

142 100

127 54.6 43.3 64.9



141 14.8 11.5 17.3




Abundance Ion 141.80 (141.50 to 142.50): V

Ion 126.80 (126.50 to 127.50): V

Ion 140.80 (140.50 to 141.50): V









#11

Tert butyl alcohol
 Concen: 213.610 ug/l
 RT: 4.72 min Scan# 928
 Delta R.T. -0.01 min
 Lab File: VN064389.D
 Acq: 30 Oct 2020 10:30

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050




Tot Ion: 59 Resp: 123915
 Ion Ratio Lower Upper
 59 100
 57 12.0 9.4 14.0




#12

1,1-Dichloroethene
 Concen: 52.421 ug/l
 RT: 3.69 min Scan# 608
 Delta R.T. -0.00 min
 Lab File: VN064389.D
 Acq: 30 Oct 2020 10:30

Tgt Ion: 96 Resp: 124541
 Ion Ratio Lower Upper
 96 100
 61 189.7 143.6 215.4
 98 63.1 48.6 73.0





#13

Acrolein

Concen: 201.061 ug/l

RT: 3.57 min Scan# 568

Delta R.T. 0.00 min

Lab File: VN064389.D

Acq: 30 Oct 2020 10:30

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050

Tot Ion: 56 Resp: 24842

Ion Ratio Lower Upper

56 100

55 75.0 62.6 94.0



#14

Allyl chloride

Concen: 48.031 ug/l

RT: 4.27 min Scan# 788

Delta R.T. -0.00 min

Lab File: VN064389.D

Acq: 30 Oct 2020 10:30




Tgt Ion: 41 Resp: 232998


Ion Ratio Lower Upper

41 100

39 81.2 61.7 92.5

76 34.2 24.0 36.0





#15

Acrylonitrile

Concen: 249.269 ug/l

RT: 4.92 min Scan# 990

Delta R.T. -0.00 min

Lab File: VN064389.D

Acq: 30 Oct 2020 10:30

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050





Tot Ion: 53 Resp: 310960

Ion Ratio Lower Upper

53 100

52 83.5 65.9 98.9

51 38.0 30.4 45.6



#16

Acetone

Concen: 299.337 ug/l

RT: 3.77 min Scan# 631

Delta R.T. -0.00 min

Lab File: VN064389.D




Acq: 30 Oct 2020 10:30


Tgt Ion: 43 Resp: 405835

Ion Ratio Lower Upper

43 100

58 26.5 22.1 33.1









#17
Carbon Disulfide
Concen: 49.508 ug/l
RT: 4.01 min Scan# 706
Delta R.T. -0.00 min
Lab File: VN064389.D
Acq: 30 Oct 2020 10:30




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tot Ion: 76 Resp: 347996
Ion Ratio Lower Upper
76 100
78 8.8 7.3 10.9



#18
Methyl Acetate
Concen: 48.787 ug/l
RT: 4.27 min Scan# 788
Delta R.T. -0.00 min
Lab File: VN064389.D
Acq: 30 Oct 2020 10:30

Tgt Ion: 43 Resp: 146451
Ion Ratio Lower Upper
43 100
74 20.9 16.7 25.1









#19

Methyl tert-butyl Ether
 Concen: 52.750 ug/l
 RT: 4.99 min Scan# 1010
 Delta R.T. -0.00 min
 Lab File: VN064389.D
 Acq: 30 Oct 2020 10:30

Instrument :
 MSVOA_N
 ClientSampleID :
 VSTDCCC050




Tot Ion: 73 Resp: 481516
 Ion Ratio Lower Upper
 73 100
 57 20.8 16.1 24.1




#20

Methylene Chloride
 Concen: 51.230 ug/l
 RT: 4.50 min Scan# 859
 Delta R.T. -0.00 min
 Lab File: VN064389.D
 Acq: 30 Oct 2020 10:30

Tgt Ion: 84 Resp: 145187
 Ion Ratio Lower Upper
 84 100
 49 133.1 103.6 155.4
 51 40.1 31.8 47.8
 86 63.4 51.3 76.9





#21

trans-1,2-Dichloroethene

Concen: 52.741 ug/l

RT: 4.99 min Scan# 1010

Delta R.T. -0.00 min

Lab File: VN064389.D

Acq: 30 Oct 2020 10:30

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050





Tot Ion: 96 Resp: 139778

Ion Ratio Lower Upper

96 100

61 149.4 128.2 192.2

98 63.7 56.1 84.1



#22

Diisopropyl ether

Concen: 54.174 ug/l

RT: 5.89 min Scan# 1292

Delta R.T. -0.01 min

Lab File: VN064389.D

Acq: 30 Oct 2020 10:30

Tgt Ion: 45 Resp: 513674




Ion Ratio Lower Upper


45 100

43 67.0 54.6 82.0

87 26.6 22.6 33.8

59 9.9 8.6 12.8





#23

Vinyl Acetate

Concen: 271.719 ug/l

RT: 5.84 min Scan# 1274

Delta R.T. -0.01 min

Lab File: VN064389.D

Acq: 30 Oct 2020 10:30

Instrument :

MSVOA_N

ClientSampleId :


VSTDCCC050

Tot Ion: 43 Resp: 2179701

Ion Ratio Lower Upper




43 100

86 9.0 7.3 10.9



Abundance Ion 43.00 (42.70 to 43.70): VN064389.D (-1183) (-)

Ion 86.00 (85.70 to 86.70): VN064389.D (-1183) (-)



#24

1,1-Dichloroethane

Concen: 52.381 ug/l

RT: 5.79 min Scan# 1261

Delta R.T. -0.00 min

Lab File: VN064389.D

Acq: 30 Oct 2020 10:30


Tgt Ion: 63 Resp: 277921

Ion Ratio Lower Upper

63 100

98 6.9 3.1 9.4



100 4.1 2.0 5.8




Abundance Ion 62.90 (62.60 to 63.60): VN064389.D (-1169) (-)

Ion 97.90 (97.60 to 98.60): VN064389.D (-1169) (-)

Ion 99.90 (99.60 to 100.60): VN064389.D (-1169) (-)





#25

2-Butanone

Concen: 263.507 ug/l

RT: 6.78 min Scan# 1567

Delta R.T. -0.00 min

Lab File: VN064389.D

Acq: 30 Oct 2020 10:30

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050

Tot Ion: 43 Resp: 528715

Ion Ratio Lower Upper

43 100

72 24.1 18.8 28.2



#26

2,2-Dichloropropane

Concen: 53.116 ug/l

RT: 6.77 min Scan# 1565

Delta R.T. -0.00 min

Lab File: VN064389.D




Acq: 30 Oct 2020 10:30


Tgt Ion: 77 Resp: 275817

Ion Ratio Lower Upper

77 100

97 20.6 10.2 30.6





#27

cis-1,2-Dichloroethene

Concen: 52.984 ug/l

RT: 6.78 min Scan# 1567

Delta R.T. -0.00 min

Lab File: VN064389.D

Acq: 30 Oct 2020 10:30

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050


Tot Ion: 96 Resp: 160705

Ion Ratio Lower Upper

96 100

61 166.9 0.0 325.0

98 66.9 0.0 128.4






Abundance Ion 95.90 (95.60 to 96.60): VN064389.D (-1474) (-)

Ion 60.90 (60.60 to 61.60): VN064389.D (-1474) (-)

Ion 97.90 (97.60 to 98.60): VN064389.D (-1474) (-)

Time-->



#28

Bromochloromethane

Concen: 49.794 ug/l

RT: 7.14 min Scan# 1680

Delta R.T. -0.01 min

Lab File: VN064389.D

Acq: 30 Oct 2020 10:30


Tgt Ion: 49 Resp: 125875

Ion Ratio Lower Upper

49 100

129 0.7 0.0 2.8

130 71.1 56.2 84.2






Abundance Ion 48.90 (48.60 to 49.60): VN064389.D (-1589) (-)

Ion 128.80 (128.50 to 129.50): VN064389.D (-1589) (-)

Ion 129.80 (129.50 to 130.50): VN064389.D (-1589) (-)

Time-->





#29

Tetrahydrofuran

Concen: 233.655 ug/l

RT: 7.16 min Scan# 1685

Delta R.T. -0.01 min

Lab File: VN064389.D

Acq: 30 Oct 2020 10:30

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050


Tot Ion: 42 Resp: 304353

Ion Ratio Lower Upper

42 100

72 39.8 31.6 47.4

71 36.7 29.4 44.0



Abundance Ion 42.00 (41.70 to 42.70): VN064320.D (-1660) (-)

Ion 72.00 (71.70 to 72.70): VN064320.D (-1660) (-)

Ion 71.00 (70.70 to 71.70): VN064320.D (-1660) (-)

7.16



150000

100000

50000

0

Time-->



#30

Chloroform

Concen: 53.362 ug/l

RT: 7.32 min Scan# 1737

Delta R.T. -0.00 min

Lab File: VN064389.D


Acq: 30 Oct 2020 10:30

Tgt Ion: 83 Resp: 302209

Ion Ratio Lower Upper

83 100

85 63.9 51.1 76.7



Abundance Ion 82.90 (82.60 to 83.60): VN064320.D (-1718) (-)

Ion 84.90 (84.60 to 85.60): VN064320.D (-1718) (-)


7.32


100000

50000

0

Time-->









#31
Cyclohexane
Concen: 47.201 ug/l
RT: 7.61 min Scan# 1826
Delta R.T. -0.00 min
Lab File: VN064389.D
Acq: 30 Oct 2020 10:30




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tot Ion: 56 Resp: 218914
Ion Ratio Lower Upper
56 100
69 34.4 25.8 38.8
84 79.0 66.3 99.5



#32
1,1,1-Trichloroethane
Concen: 52.451 ug/l
RT: 7.52 min Scan# 1799
Delta R.T. -0.00 min
Lab File: VN064389.D
Acq: 30 Oct 2020 10:30

Tgt Ion: 97 Resp: 285242
Ion Ratio Lower Upper
97 100
99 63.0 50.6 76.0
61 45.4 35.8 53.6









#33
 1,2-Dichloroethane-d4
 Concen: 47.441 ug/l
 RT: 7.98 min Scan# 1941
 Delta R.T. -0.01 min
 Lab File: VN064389.D
 Acq: 30 Oct 2020 10:30




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050


Tot Ion: 65 Resp: 189579
 Ion Ratio Lower Upper
 65 100
 67 49.4 0.0 99.2



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.55 min Scan# 2117
 Delta R.T. -0.00 min
 Lab File: VN064389.D
 Acq: 30 Oct 2020 10:30

Tgt Ion: 114 Resp: 396204
 Ion Ratio Lower Upper
 114 100
 63 22.3 0.0 44.6
 88 17.3 0.0 33.4









#35
 Dibromofluoromethane
 Concen: 49.278 ug/l
 RT: 7.54 min Scan# 1805
 Delta R.T. -0.00 min
 Lab File: VN064389.D
 Acq: 30 Oct 2020 10:30




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050


Tot Ion:113 Resp: 130202
 Ion Ratio Lower Upper
 113 100
 111 103.2 81.5 122.3
 192 18.5 15.0 22.4



#36
 1,1-Dichloropropene
 Concen: 54.503 ug/l
 RT: 7.75 min Scan# 1868
 Delta R.T. -0.01 min
 Lab File: VN064389.D
 Acq: 30 Oct 2020 10:30

Tgt Ion: 75 Resp: 217937
 Ion Ratio Lower Upper
 75 100
 110 33.8 16.8 50.3
 77 31.1 24.2 36.4









#37
Ethyl Acetate
Concen: 50.072 ug/l
RT: 6.87 min Scan# 1597
Delta R.T. -0.01 min
Lab File: VN064389.D
Acq: 30 Oct 2020 10:30




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tot Ion: 43 Resp: 202240
Ion Ratio Lower Upper
43 100
61 12.5 10.0 15.0
70 10.0 7.1 10.7



#38
Carbon Tetrachloride
Concen: 55.711 ug/l
RT: 7.73 min Scan# 1863
Delta R.T. -0.00 min
Lab File: VN064389.D
Acq: 30 Oct 2020 10:30

Tgt Ion: 117 Resp: 258263
Ion Ratio Lower Upper
117 100
119 93.7 73.7 110.5
121 30.7 22.8 34.2









#39
Methvlcvclohexane
Concen: 47.014 ug/l
RT: 9.04 min Scan# 2271
Delta R.T. -0.00 min
Lab File: VN064389.D
Acq: 30 Oct 2020 10:30




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tot Ion: 83 Resp: 186402
Ion Ratio Lower Upper
83 100
55 80.4 66.2 99.4
98 45.4 36.0 54.0



#40
Benzene
Concen: 53.766 ug/l
RT: 8.00 min Scan# 1947
Delta R.T. -0.00 min
Lab File: VN064389.D
Acq: 30 Oct 2020 10:30

Tgt Ion: 78 Resp: 598429
Ion Ratio Lower Upper
78 100
77 22.2 18.8 28.2






#41
Methacrylonitrile
Concen: 50.337 ug/l
RT: 7.12 min Scan# 1672
Delta R.T. -0.01 min
Lab File: VN064389.D
Acq: 30 Oct 2020 10:30

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tot Ion:	41	Resp:	100924
Ion	Ratio	Lower	Upper
41	100		
39	51.5	48.6	72.8
67	72.0	61.4	92.2
52	31.3	27.6	41.4






Abundance

Ion 41.00 (40.70 to 41.70): VN0

Ion 39.00 (38.70 to 39.70): VN0


Ion 67.00 (66.70 to 67.70): VN0

Ion 52.00 (51.70 to 52.70): VN0



#42
1,2-Dichloroethane
Concen: 55.608 ug/l
RT: 8.08 min Scan# 1972
Delta R.T. -0.00 min
Lab File: VN064389.D
Acq: 30 Oct 2020 10:30



Tgt Ion:	62	Resp:	267363
Ion	Ratio	Lower	Upper
62	100		
98	6.8	0.0	14.2




Abundance

Ion 61.90 (61.60 to 62.60): VN0

Ion 97.90 (97.60 to 98.60): VN0





#43

Isopropyl Acetate

Concen: 53.225 ug/l

RT: 8.12 min Scan# 1984

Delta R.T. -0.01 min

Lab File: VN064389.D

Acq: 30 Oct 2020 10:30

Instrument :


MSVOA_N

ClientSampleId :

VSTDCCC050

Tot Ion: 43 Resp: 367481

Ion	Ratio	Lower	Upper
43	100		
61	17.2	19.1	28.7#
87	11.7	9.3	13.9




Abundance Ion 43.00 (42.70 to 43.70): VN064320.D (-1967) (-)

Ion 61.00 (60.70 to 61.70): VN064320.D (-1967) (-)

Ion 87.00 (86.70 to 87.70): VN064320.D (-1967) (-)

Time-->




Abundance

Ion 43.00 (42.70 to 43.70): VN064320.D (-1967) (-)

Ion 61.00 (60.70 to 61.70): VN064320.D (-1967) (-)

Ion 87.00 (86.70 to 87.70): VN064320.D (-1967) (-)

Time-->



#44

Trichloroethene

Concen: 55.481 ug/l

RT: 8.80 min Scan# 2195


Delta R.T. -0.00 min

Lab File: VN064389.D

Acq: 30 Oct 2020 10:30

Tgt Ion: 130 Resp: 161767


Ion	Ratio	Lower	Upper
130	100		
95	110.1	0.0	203.6



Abundance Ion 129.80 (129.50 to 130.50): VN064320.D (-2180) (-)

Ion 94.90 (94.60 to 95.60): VN064320.D (-2180) (-)

Time-->




Abundance

Ion 129.80 (129.50 to 130.50): VN064320.D (-2180) (-)

Ion 94.90 (94.60 to 95.60): VN064320.D (-2180) (-)





Time-->



#45
 1,2-Dichloropropane
 Concen: 53.708 ug/l
 RT: 9.08 min Scan# 2283
 Delta R.T. -0.00 min
 Lab File: VN064389.D
 Acq: 30 Oct 2020 10:30




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050


Tot Ion: 63 Resp: 160077
 Ion Ratio Lower Upper
 63 100
 65 28.6 24.3 36.5



#46
 Dibromomethane
 Concen: 54.329 ug/l
 RT: 9.17 min Scan# 2311
 Delta R.T. -0.00 min
 Lab File: VN064389.D
 Acq: 30 Oct 2020 10:30

Tgt Ion: 93 Resp: 113282
 Ion Ratio Lower Upper
 93 100
 95 83.5 64.4 96.6
 174 92.5 71.4 107.0





#47

Bromodichloromethane

Concen: 53.337 ug/l

RT: 9.37 min Scan# 2372

Delta R.T. -0.00 min

Lab File: VN064389.D

Acq: 30 Oct 2020 10:30

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050


Tot Ion: 83 Resp: 246469

Ion Ratio Lower Upper

83 100

85 63.9 52.1 78.1

127 7.7 6.7 10.1



Abundance Ion 82.90 (82.60 to 83.60): VN064389.D (-2280) (-)

Ion 84.90 (84.60 to 85.60): VN064389.D (-2280) (-)

Ion 126.80 (126.50 to 127.50): VN064389.D (-2280) (-)

9.37

150000



100000

50000

0

9.30 9.35 9.40 9.45

Time-->



#48

Methyl methacrylate

Concen: 49.615 ug/l

RT: 9.16 min Scan# 2308

Delta R.T. -0.00 min

Lab File: VN064389.D

Acq: 30 Oct 2020 10:30


Tgt Ion: 41 Resp: 176738

Ion Ratio Lower Upper

41 100

69 74.2 59.1 88.7

39 65.6 51.3 76.9



Abundance Ion 41.00 (40.70 to 41.70): VN064389.D (-2216) (-)

Ion 69.00 (68.70 to 69.70): VN064389.D (-2216) (-)

Ion 39.00 (38.70 to 39.70): VN064389.D (-2216) (-)

9.16

120000

100000

80000

60000


40000


20000

0

9.10 9.15 9.20

Time-->









#49
1,4-Dioxane
Concen: 941.060 ug/l
RT: 9.16 min Scan# 2308
Delta R.T. -0.00 min
Lab File: VN064389.D
Acq: 30 Oct 2020 10:30

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tot Ion: 88 Resp: 49410
Ion Ratio Lower Upper
88 100
43 41.5 31.8 47.8
58 71.9 60.3 90.5




Abundance Ion 88.00 (87.70 to 88.70): VN0
Ion 43.00 (42.70 to 43.70): VN0
Ion 58.00 (57.70 to 58.70): VN0






#50
Toluene-d8
Concen: 48.891 ug/l
RT: 10.05 min Scan# 2586
Delta R.T. -0.00 min
Lab File: VN064389.D
Acq: 30 Oct 2020 10:30

Tgt Ion: 98 Resp: 477814
Ion Ratio Lower Upper
98 100
100 64.4 50.9 76.3



Abundance Ion 98.00 (97.70 to 98.70): VN0
Ion 100.00 (99.70 to 100.70): VN0





#51

4-Methyl-2-Pentanone

Concen: 256.209 ug/l

RT: 9.95 min Scan# 2553

Delta R.T. -0.00 min

Lab File: VN064389.D

Acq: 30 Oct 2020 10:30

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050

Tot Ion: 43 Resp: 1028266

Ion Ratio Lower Upper

43 100

58 35.2 28.6 42.8



#52

Toluene

Concen: 54.690 ug/l

RT: 10.12 min Scan# 2607

Delta R.T. -0.00 min

Lab File: VN064389.D




Acq: 30 Oct 2020 10:30


Tgt Ion: 92 Resp: 384111

Ion Ratio Lower Upper

92 100

91 175.0 139.9 209.9





#53

t-1,3-Dichloropropene

Concen: 56.227 ug/l

RT: 10.35 min Scan# 2677

Delta R.T. -0.00 min

Lab File: VN064389.D

Acq: 30 Oct 2020 10:30

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050

Tot Ion: 75 Resp: 274725

Ion Ratio Lower Upper

75 100

77 31.1 24.8 37.2



#54

cis-1,3-Dichloropropene

Concen: 55.233 ug/l

RT: 9.80 min Scan# 2508

Delta R.T. -0.00 min

Lab File: VN064389.D

Acq: 30 Oct 2020 10:30




Tgt Ion: 75 Resp: 277796


Ion Ratio Lower Upper

75 100

77 31.1 25.0 37.4

39 58.6 45.5 68.3









#55
 1,1,2-Trichloroethane
 Concen: 54.578 ug/l
 RT: 10.53 min Scan# 2734
 Delta R.T. -0.00 min
 Lab File: VN064389.D
 Acq: 30 Oct 2020 10:30

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050

Tot Ion: 97 Resp: 151310




Ion	Ratio	Lower	Upper
97	100		
83	87.2	68.9	103.3
85	55.2	43.6	65.4
99	60.7	49.2	73.8




#56
 Ethyl methacrylate
 Concen: 53.482 ug/l
 RT: 10.40 min Scan# 2692
 Delta R.T. -0.00 min
 Lab File: VN064389.D
 Acq: 30 Oct 2020 10:30

Tgt Ion: 69 Resp: 224449

Ion	Ratio	Lower	Upper
69	100		
41	76.3	61.4	92.2
39	47.0	36.5	54.7





#57

1,3-Dichloropropane

Concen: 55.405 ug/l

RT: 10.68 min Scan# 2779

Delta R.T. -0.00 min

Lab File: VN064389.D

Acq: 30 Oct 2020 10:30

Instrument :

MSVOA_N

ClientSampleId :


VSTDCCC050

Tot Ion: 76 Resp: 262023

Ion Ratio Lower Upper




76 100

78 33.0 25.8 38.6



Abundance Ion 75.90 (75.60 to 76.60): VN0

Ion 77.90 (77.60 to 78.60): VN0



#58

2-Chloroethyl Vinyl ether

Concen: 252.967 ug/l

RT: 9.66 min Scan# 2463

Delta R.T. -0.00 min

Lab File: VN064389.D


Acq: 30 Oct 2020 10:30

Tgt Ion: 63 Resp: 579324

Ion Ratio Lower Upper



63 100


106 26.5 21.0 31.4



Abundance Ion 62.90 (62.60 to 63.60): VN0

Ion 105.90 (105.60 to 106.60): V









#59
2-Hexanone
Concen: 266.019 ug/l
RT: 10.72 min Scan# 2793
Delta R.T. -0.00 min
Lab File: VN064389.D
Acq: 30 Oct 2020 10:30




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tot Ion: 43 Resp: 786313
Ion Ratio Lower Upper
43 100
58 48.9 25.0 75.0



#60
Dibromochloromethane
Concen: 56.775 ug/l
RT: 10.87 min Scan# 2840
Delta R.T. -0.00 min
Lab File: VN064389.D
Acq: 30 Oct 2020 10:30

Tgt Ion: 129 Resp: 186104
Ion Ratio Lower Upper
129 100
127 77.4 38.5 115.3









#61
 1,2-Dibromoethane
 Concen: 56.621 ug/l
 RT: 10.97 min Scan# 2872
 Delta R.T. -0.00 min
 Lab File: VN064389.D
 Acq: 30 Oct 2020 10:30




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050


Tot Ion: 107 Resp: 159946
 Ion Ratio Lower Upper
 107 100
 109 93.0 75.8 113.8



#62
 4-Bromofluorobenzene
 Concen: 48.317 ug/l
 RT: 12.37 min Scan# 3307
 Delta R.T. -0.00 min
 Lab File: VN064389.D
 Acq: 30 Oct 2020 10:30

Tgt Ion: 95 Resp: 184112
 Ion Ratio Lower Upper
 95 100
 174 77.8 0.0 147.4
 176 74.0 0.0 143.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.38 min Scan# 2997
Delta R.T. -0.00 min
Lab File: VN064389.D
Acq: 30 Oct 2020 10:30




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tot Ion:117 Resp: 376925
Ion Ratio Lower Upper
117 100
82 57.2 48.8 73.2
119 31.1 25.1 37.7



#64
Tetrachloroethene
Concen: 49.881 ug/l
RT: 10.60 min Scan# 2755
Delta R.T. -0.00 min
Lab File: VN064389.D
Acq: 30 Oct 2020 10:30

Tgt Ion:164 Resp: 154529
Ion Ratio Lower Upper
164 100
166 126.6 100.7 151.1
129 99.4 83.7 125.5
131 96.5 78.2 117.2





#65

Chlorobenzene

Concen: 54.057 ug/l

RT: 11.40 min Scan# 3005

Delta R.T. -0.00 min

Lab File: VN064389.D

Acq: 30 Oct 2020 10:30

Instrument :

MSVOA_N

ClientSampleId :


VSTDCCC050

Tot Ion:112 Resp: 413076

Ion Ratio Lower Upper




112 100

114 31.5 25.4 38.0



Abundance Ion 111.90 (111.60 to 112.60): V

Ion 113.90 (113.60 to 114.60): V



#66

1,1,1,2-Tetrachloroethane

Concen: 53.282 ug/l

RT: 11.48 min Scan# 3029

Delta R.T. -0.00 min

Lab File: VN064389.D

Acq: 30 Oct 2020 10:30


Tgt Ion:131 Resp: 164565

Ion Ratio Lower Upper

131 100

133 95.9 49.3 147.8



119 62.7 32.3 96.9




Abundance Ion 130.80 (130.50 to 131.50): V

Ion 132.80 (132.50 to 133.50): V

Ion 118.80 (118.50 to 119.50): V









#67
Ethyl Benzene
Concen: 53.015 ug/l
RT: 11.48 min Scan# 3030
Delta R.T. -0.00 min
Lab File: VN064389.D
Acq: 30 Oct 2020 10:30




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tot Ion: 91 Resp: 753168
Ion Ratio Lower Upper
91 100
106 31.6 23.7 35.5



#68
m/p-Xylenes
Concen: 105.804 ug/l
RT: 11.59 min Scan# 3064
Delta R.T. -0.00 min
Lab File: VN064389.D
Acq: 30 Oct 2020 10:30

Tgt Ion: 106 Resp: 557044
Ion Ratio Lower Upper
106 100
91 213.2 170.2 255.4









#69
o-Xylene
Concen: 53.231 ug/l
RT: 11.92 min Scan# 3166
Delta R.T. -0.00 min
Lab File: VN064389.D
Acq: 30 Oct 2020 10:30




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tot Ion:106 Resp: 270002
Ion Ratio Lower Upper
106 100
91 224.6 114.5 343.5



#70
Styrene
Concen: 54.972 ug/l
RT: 11.94 min Scan# 3171
Delta R.T. -0.00 min
Lab File: VN064389.D
Acq: 30 Oct 2020 10:30

Tgt Ion:104 Resp: 457173
Ion Ratio Lower Upper
104 100
78 55.5 43.7 65.5
103 54.5 43.3 64.9





#71

Bromoform

Concen: 51.504 ug/l

RT: 12.10 min Scan# 3221

Delta R.T. -0.00 min

Lab File: VN064389.D

Acq: 30 Oct 2020 10:30

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050


Tot Ion:173 Resp: 124850

Ion Ratio Lower Upper

173 100

175 50.8 25.4 76.0




254 0.0 0.0 0.0



Abundance Ion 172.70 (172.40 to 173.40): V

Ion 174.70 (174.40 to 175.40): V

Ion 254.40 (254.10 to 255.10): V



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.32 min Scan# 3600

Delta R.T. -0.00 min

Lab File: VN064389.D

Acq: 30 Oct 2020 10:30


Tgt Ion:152 Resp: 174259

Ion Ratio Lower Upper

152 100

115 62.8 31.5 94.5



150 174.6 0.0 346.2




Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V





#73

Isopropylbenzene

Concen: 53.522 ug/l

RT: 12.22 min Scan# 3259

Delta R.T. -0.00 min

Lab File: VN064389.D

Acq: 30 Oct 2020 10:30

Instrument :

MSVOA_N

ClientSampleId :


VSTDCCC050

Tot Ion:105 Resp: 698539

Ion Ratio Lower Upper

105 100

120 25.8 13.2 39.5



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V

12.22

400000



300000

200000

100000

0

Time--> 12.15 12.20 12.25 12.30



#74

N-ethyl acetate

Concen: 55.473 ug/l

RT: 12.04 min Scan# 3204

Delta R.T. -0.00 min

Lab File: VN064389.D

Acq: 30 Oct 2020 10:30

Tgt Ion: 43 Resp: 318677


Ion Ratio Lower Upper

43 100

70 37.9 30.6 45.8

55 22.6 18.2 27.4

61 21.2 17.6 26.4



Abundance Ion 43.00 (42.70 to 43.70): V

Ion 70.00 (69.70 to 70.70): V

Ion 55.00 (54.70 to 55.70): V

Ion 61.00 (60.70 to 61.70): V

300000

250000

200000


150000


100000

50000

0

Time--> 11.95 12.00 12.05 12.10





#75

1,1,2,2-Tetrachloroethane

Concen: 52.607 ug/l

RT: 12.48 min Scan# 3339

Delta R.T. -0.00 min

Lab File: VN064389.D

Acq: 30 Oct 2020 10:30

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050


Tot Ion: 83 Resp: 199575

Ion Ratio Lower Upper

83 100

131 10.9 5.4 16.2




85 64.3 33.0 98.9



Abundance Ion 82.90 (82.60 to 83.60): VN064320.D (-3325) (-)

Ion 130.80 (130.50 to 131.50): VN064320.D (-3325) (-)

Ion 84.90 (84.60 to 85.60): VN064320.D (-3325) (-)



#76

1,2,3-Trichloropropane

Concen: 70.540 ug/l

RT: 12.52 min Scan# 3354

Delta R.T. -0.00 min

Lab File: VN064389.D


Acq: 30 Oct 2020 10:30

Tgt Ion: 75 Resp: 294366

Ion Ratio Lower Upper

75 100

77 0.0 0.0 0.0



Abundance Ion 74.90 (74.60 to 75.60): VN064320.D (-3349) (-)

Ion 77.00 (76.70 to 77.70): VN064320.D (-3349) (-)

Ion 77.00 (76.70 to 77.70): VN064320.D (-3349) (-)

Ion 77.00 (76.70 to 77.70): VN064320.D (-3349) (-)

Ion 77.00 (76.70 to 77.70): VN064320.D (-3349) (-)

Ion 77.00 (76.70 to 77.70): VN064320.D (-3349) (-)

Ion 77.00 (76.70 to 77.70): VN064320.D (-3349) (-)

Ion 77.00 (76.70 to 77.70): VN064320.D (-3349) (-)

Ion 77.00 (76.70 to 77.70): VN064320.D (-3349) (-)

Ion 77.00 (76.70 to 77.70): VN064320.D (-3349) (-)

Ion 77.00 (76.70 to 77.70): VN064320.D (-3349) (-)

Ion 77.00 (76.70 to 77.70): VN064320.D (-3349) (-)

Ion 77.00 (76.70 to 77.70): VN064320.D (-3349) (-)

Ion 77.00 (76.70 to 77.70): VN064320.D (-3349) (-)

Ion 77.00 (76.70 to 77.70): VN064320.D (-3349) (-)

Ion 77.00 (76.70 to 77.70): VN064320.D (-3349) (-)

Ion 77.00 (76.70 to 77.70): VN064320.D (-3349) (-)

Ion 77.00 (76.70 to 77.70): VN064320.D (-3349) (-)

Ion 77.00 (76.70 to 77.70): VN064320.D (-3349) (-)

Ion 77.00 (76.70 to 77.70): VN064320.D (-3349) (-)

Ion 77.00 (76.70 to 77.70): VN064320.D (-3349) (-)

Ion 77.00 (76.70 to 77.70): VN064320.D (-3349) (-)

Ion 77.00 (76.70 to 77.70): VN064320.D (-3349) (-)

Ion 77.00 (76.70 to 77.70): VN064320.D (-3349) (-)

Ion 77.00 (76.70 to 77.70): VN064320.D (-3349) (-)

Ion 77.00 (76.70 to 77.70): VN064320.D (-3349) (-)

Ion 77.00 (76.70 to 77.70): VN064320.D (-3349) (-)



Ion 77.00 (76.70 to 77.70): VN064320.D (-3349) (-)


Ion 77.00 (76.70 to 77.70): VN064320.D (-3349) (-)

Ion 77.00 (76.70 to 77.70): VN064320.D (-3349) (-)

Ion 77.00 (76.70 to 77.70): VN064320.D (-3349) (-)

Ion 77.00 (76.70 to 77.70): VN064320.D (-3349) (-)





#77

Bromobenzene

Concen: 54.789 ug/l

RT: 12.50 min Scan# 3346

Delta R.T. -0.00 min

Lab File: VN064389.D

Acq: 30 Oct 2020 10:30

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050


Tot Ion:156 Resp: 175103

Ion Ratio Lower Upper

156 100

77 222.3 111.9 335.7




158 99.5 49.0 147.0



Abundance Ion 155.80 (155.50 to 156.50): V

Ion 77.00 (76.70 to 77.70): VN0

Ion 157.80 (157.50 to 158.50): V



#78

n-propylbenzene

Concen: 53.593 ug/l

RT: 12.56 min Scan# 3366

Delta R.T. -0.00 min

Lab File: VN064389.D


Acq: 30 Oct 2020 10:30

Tgt Ion: 91 Resp: 791891

Ion Ratio Lower Upper



91 100


120 21.7 11.2 33.5



Abundance Ion 91.00 (90.70 to 91.70): VN0

Ion 120.00 (119.70 to 120.70): V









#79
 2-Chlorotoluene
 Concen: 52.943 ug/l
 RT: 12.65 min Scan# 3392
 Delta R.T. -0.00 min
 Lab File: VN064389.D
 Acq: 30 Oct 2020 10:30




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050


Tot Ion: 91 Resp: 500073
 Ion Ratio Lower Upper
 91 100
 126 32.5 16.2 48.5



#80
 1,3,5-Trimethylbenzene
 Concen: 52.407 ug/l
 RT: 12.70 min Scan# 3410
 Delta R.T. -0.00 min
 Lab File: VN064389.D
 Acq: 30 Oct 2020 10:30

Tgt Ion: 105 Resp: 578757
 Ion Ratio Lower Upper
 105 100
 120 46.8 22.9 68.8





#81

trans-1,4-Dichloro-2-butene

Concen: 52.697 ug/l

RT: 12.27 min Scan# 3276

Delta R.T. -0.00 min

Lab File: VN064389.D

Acq: 30 Oct 2020 10:30

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050





Tot Ion: 75 Resp: 76343

Ion Ratio Lower Upper

75 100

53 105.8 81.8 122.8

89 45.8 39.9 59.9



#82

4-Chlorotoluene

Concen: 55.433 ug/l

RT: 12.74 min Scan# 3422

Delta R.T. -0.00 min

Lab File: VN064389.D




Acq: 30 Oct 2020 10:30


Tgt Ion: 91 Resp: 533399

Ion Ratio Lower Upper

91 100

126 30.8 15.8 47.3









#83
 tert-Butylbenzene
 Concen: 50.699 ug/l
 RT: 12.96 min Scan# 3491
 Delta R.T. -0.00 min
 Lab File: VN064389.D
 Acq: 30 Oct 2020 10:30




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050


Tot Ion:119 Resp: 459444
 Ion Ratio Lower Upper
 119 100
 91 72.3 36.4 109.1
 134 24.4 12.0 36.1



#84
 1,2,4-Trimethylbenzene
 Concen: 53.135 ug/l
 RT: 13.01 min Scan# 3505
 Delta R.T. -0.00 min
 Lab File: VN064389.D
 Acq: 30 Oct 2020 10:30

Tgt Ion:105 Resp: 586330
 Ion Ratio Lower Upper
 105 100
 120 43.6 21.3 64.0









#85
 sec-Butylbenzene
 Concen: 48.328 ug/l
 RT: 13.14 min Scan# 3547
 Delta R.T. -0.00 min
 Lab File: VN064389.D
 Acq: 30 Oct 2020 10:30




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050


Tot Ion:105 Resp: 566741
 Ion Ratio Lower Upper
 105 100
 134 19.3 9.6 28.8



#86
 p-Isopropyltoluene
 Concen: 49.612 ug/l
 RT: 13.26 min Scan# 3583
 Delta R.T. -0.00 min
 Lab File: VN064389.D
 Acq: 30 Oct 2020 10:30

Tgt Ion:119 Resp: 524144
 Ion Ratio Lower Upper
 119 100
 134 25.8 12.8 38.4
 91 26.5 13.0 38.9









#87
 1,3-Dichlorobenzene
 Concen: 53.912 ug/l
 RT: 13.25 min Scan# 3581
 Delta R.T. -0.00 min
 Lab File: VN064389.D
 Acq: 30 Oct 2020 10:30




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050


Tot Ion:146 Resp: 301633
 Ion Ratio Lower Upper
 146 100
 111 43.3 21.9 65.5
 148 63.5 32.6 98.0



#88
 1,4-Dichlorobenzene
 Concen: 53.101 ug/l
 RT: 13.33 min Scan# 3606
 Delta R.T. -0.00 min
 Lab File: VN064389.D
 Acq: 30 Oct 2020 10:30

Tgt Ion:146 Resp: 302593
 Ion Ratio Lower Upper
 146 100
 111 42.7 21.8 65.4
 148 64.0 31.8 95.3









#89
n-Butylbenzene
Concen: 47.754 ug/l
RT: 13.59 min Scan# 3684
Delta R.T. -0.00 min
Lab File: VN064389.D
Acq: 30 Oct 2020 10:30




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tot Ion: 91 Resp: 444453
Ion Ratio Lower Upper
91 100
92 49.7 26.2 78.5
134 24.2 12.4 37.4



#90
Hexachloroethane
Concen: 48.825 ug/l
RT: 13.85 min Scan# 3765
Delta R.T. -0.00 min
Lab File: VN064389.D
Acq: 30 Oct 2020 10:30

Tgt Ion: 117 Resp: 96875
Ion Ratio Lower Upper
117 100
201 81.3 39.8 119.3









#91
 1,2-Dichlorobenzene
 Concen: 53.298 ug/l
 RT: 13.62 min Scan# 3696
 Delta R.T. -0.00 min
 Lab File: VN064389.D
 Acq: 30 Oct 2020 10:30




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050


Tot Ion	Ratio	Lower	Upper
146	100		
111	44.0	23.0	69.0
148	63.3	32.1	96.3



#92
 1,2-Dibromo-3-Chloropropane
 Concen: 50.408 ug/l
 RT: 14.24 min Scan# 3888
 Delta R.T. -0.00 min
 Lab File: VN064389.D
 Acq: 30 Oct 2020 10:30

Tgt Ion	Ratio	Lower	Upper
75	100		
155	70.5	35.1	105.5
157	88.4	43.2	129.6









#93
 1,2,4-Trichlorobenzene
 Concen: 45.559 ug/l
 RT: 14.88 min Scan# 4087
 Delta R.T. 0.00 min
 Lab File: VN064389.D
 Acq: 30 Oct 2020 10:30




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050


Tot Ion	Ratio	Lower	Upper
180	100		
182	95.6	46.6	139.7
145	31.6	15.3	45.9



#94
 Hexachlorobutadiene
 Concen: 40.952 ug/l
 RT: 14.98 min Scan# 4118
 Delta R.T. -0.00 min
 Lab File: VN064389.D
 Acq: 30 Oct 2020 10:30

Tgt Ion	Ratio	Lower	Upper
225	100		
223	63.2	31.3	93.8
227	62.9	32.3	96.9





#95

Naphthalene

Concen: 43.472 ug/l

RT: 15.10 min Scan# 4155

Delta R.T. 0.00 min

Lab File: VN064389.D

Acq: 30 Oct 2020 10:30

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050


Tot Ion:128 Resp: 425018

Ion Ratio Lower Upper

128 100

127 13.2 10.4 15.6




129 11.3 8.6 13.0



Abundance Ion 128.00 (127.70 to 128.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 129.00 (128.70 to 129.70): V



#96

1,2,3-Trichlorobenzene

Concen: 43.978 ug/l

RT: 15.28 min Scan# 4211

Delta R.T. -0.00 min

Lab File: VN064389.D

Acq: 30 Oct 2020 10:30


Tgt Ion:180 Resp: 134673

Ion Ratio Lower Upper

180 100

182 98.6 46.1 138.3

145 32.8 16.2 48.5



Abundance Ion 179.80 (179.50 to 180.50): V

Ion 181.80 (181.50 to 182.50): V

Ion 144.90 (144.60 to 145.60): V

