

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN103020\
 Data File : VN064390.D
 Acq On : 30 Oct 2020 11:03
 Operator : JC/MD
 Sample : VN1030MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1030MBL01

Quant Time: Oct 30 13:44:20 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N102720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 27 12:23:17 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	209416	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	369472	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	360771	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	144311	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.98	65	169503	50.37	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.74%	
35) Dibromofluoromethane	7.54	113	121451	49.29	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.58%	
50) Toluene-d8	10.05	98	463058	50.81	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.62%	
62) 4-Bromofluorobenzene	12.37	95	182454	51.35	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.70%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN103020\
Data File : VN064390.D
Acq On : 30 Oct 2020 11:03
Operator : JC/MD
Sample : VN1030MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1030MBL01

Quant Time: Oct 30 13:44:20 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N102720W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 27 12:23:17 2020
Response via : Initial Calibration

