

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN103020\
 Data File : VN064399.D
 Acq On : 30 Oct 2020 14:54
 Operator : JC/MD
 Sample : L4560-03
 Misc : 5.59g/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 B308-B309-LOC2-SO-0003

Quant Time: Oct 30 14:45:01 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N102720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 27 12:23:17 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	207676	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	370461	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	367774	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	164356	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	7.98	65	172731	51.76	ug/l	0.00
Spiked Amount			Recovery	=	103.52%	
35) Dibromofluoromethane	7.54	113	121024	48.99	ug/l	0.00
Spiked Amount			Recovery	=	97.98%	
50) Toluene-d8	10.06	98	464836	50.87	ug/l	0.00
Spiked Amount			Recovery	=	101.74%	
62) 4-Bromofluorobenzene	12.37	95	200144	56.17	ug/l	0.00
Spiked Amount			Recovery	=	112.34%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN103020\
Data File : VN064399.D
Acq On : 30 Oct 2020 14:54
Operator : JC/MD
Sample : L4560-03
Misc : 5.59g/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
B308-B309-LOC2-SO-0003

Quant Time: Oct 30 14:45:01 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N102720W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 27 12:23:17 2020
Response via : Initial Calibration

