

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN103020\
 Data File : VN064405.D
 Acq On : 30 Oct 2020 17:16
 Operator : JC/MD
 Sample : PB132725ZHE#09
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB132725ZHE#09

Manual Integrations
 APPROVED

MMDadoda
 11/2/2020 4:12:06 PM

Quant Time: Nov 02 03:09:30 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N102720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 27 12:23:17 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	209099	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	367047	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	372583	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	150219	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.98	65	172445	51.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.64%	
35) Dibromofluoromethane	7.55	113	122614	50.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.18%	
50) Toluene-d8	10.06	98	468096	51.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.40%	
62) 4-Bromofluorobenzene	12.38	95	195024	55.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.50%	
Target Compounds						
18) Methyl Acetate	4.29	43	3982m	1.578	ug/l	
20) Methylene Chloride	4.50	84	6300	2.644	ug/l #	89
43) Isopropyl Acetate	8.12	43	77613	12.134	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN103020\
Data File : VN064405.D
Acq On : 30 Oct 2020 17:16
Operator : JC/MD
Sample : PB132725ZHE#09
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB132725ZHE#09

Manual Integrations
APPROVED

MMDadoda
11/2/2020 4:12:06 PM

Quant Time: Nov 02 03:09:30 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N102720W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 27 12:23:17 2020
Response via : Initial Calibration

