

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN103023\
 Data File : VN079726.D
 Acq On : 30 Oct 2023 16:24
 Operator : JC\MD
 Sample : VN1030WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1030WBL01

Quant Time: Oct 31 07:30:39 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103023W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 31 07:23:24 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	250437	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	457168	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	389253	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	100696	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	225906	56.171	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	112.340%
35) Dibromofluoromethane	8.171	113	175202	53.633	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.260%
50) Toluene-d8	10.571	98	601766	49.415	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.840%
62) 4-Bromofluorobenzene	12.853	95	150795	39.322	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	78.640%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN103023\
Data File : VN079726.D
Acq On : 30 Oct 2023 16:24
Operator : JC\MD
Sample : VN1030WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1030WBL01

Quant Time: Oct 31 07:30:39 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103023W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 31 07:23:24 2023
Response via : Initial Calibration

