

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN103023\
 Data File : VN079730.D
 Acq On : 30 Oct 2023 18:18
 Operator : JC\MD
 Sample : PB156735TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB156735TB

Quant Time: Oct 31 07:33:04 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103023W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 31 07:23:24 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	256156	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	480089	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	435179	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	111865	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	215725	52.442	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 104.880%	
35) Dibromofluoromethane	8.171	113	168080	48.996	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 98.000%	
50) Toluene-d8	10.571	98	634151	49.588	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 99.180%	
62) 4-Bromofluorobenzene	12.853	95	160831	39.936	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 79.880%#	
Target Compounds						
						Qvalue
16) Acetone	4.436	43	27222	15.468	ug/l #	85
20) Methylene Chloride	5.289	84	13283	3.811	ug/l #	87
43) Isopropyl Acetate	8.694	43	180983	22.442	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN103023\
Data File : VN079730.D
Acq On : 30 Oct 2023 18:18
Operator : JC\MD
Sample : PB156735TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB156735TB

Quant Time: Oct 31 07:33:04 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103023W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 31 07:23:24 2023
Response via : Initial Calibration

