

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN103023\
 Data File : VN079732.D
 Acq On : 30 Oct 2023 19:06
 Operator : JC\MD
 Sample : 05138-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 348

Quant Time: Oct 31 07:33:58 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103023W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 31 07:23:24 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	243427	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	467718	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	423359	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	108627	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	234451	59.975	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	119.940%
35) Dibromofluoromethane	8.171	113	180494	54.006	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	108.020%
50) Toluene-d8	10.571	98	647394	51.963	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.920%
62) 4-Bromofluorobenzene	12.853	95	164830	42.012	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	84.020%
Target Compounds						
						Qvalue
16) Acetone	4.442	43	47840	28.604	ug/l #	88
20) Methylene Chloride	5.271	84	15483	4.675	ug/l #	89
43) Isopropyl Acetate	8.694	43	158142	20.129	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN103023\
Data File : VN079732.D
Acq On : 30 Oct 2023 19:06
Operator : JC\MD
Sample : 05138-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
348

Quant Time: Oct 31 07:33:58 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103023W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 31 07:23:24 2023
Response via : Initial Calibration

