

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN103118\
 Data File : VN052083.D
 Acq On : 31 Oct 2018 19:26
 Operator : MD\SY
 Sample : J5202-08
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PL-02-103018-A

Manual Integrations
 APPROVED

MMDadoda
 11/2/2018 1:11:29 PM

Quant Time: Nov 01 02:27:24 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N103018W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 31 02:45:22 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	285855	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	524891	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	505521	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	204868	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	244749	53.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.72%	
35) Dibromofluoromethane	7.59	113	200625	61.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	122.20%	
50) Toluene-d8	10.09	98	791411	60.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	120.60%	
62) 4-Bromofluorobenzene	12.40	95	271639	56.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.76%	
Target Compounds						
16) Acetone	3.82	43	11936	12.19	ug/l	# 87
18) Methyl Acetate	4.33	43	4891m	1.57	ug/l	
20) Methylene Chloride	4.53	84	35320075m	9466.85	ug/l	
43) Isopropyl Acetate	8.17	43	161901	25.48	ug/l	97
95) Naphthalene	15.13	128	9110	1.55	ug/l	# 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN103118\
Data File : VN052083.D
Acq On : 31 Oct 2018 19:26
Operator : MD\SY
Sample : J5202-08
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PL-02-103018-A

Manual Integrations
APPROVED

MMDadoda
11/2/2018 1:11:29 PM

Quant Time: Nov 01 02:27:24 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N103018W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 31 02:45:22 2018
Response via : Initial Calibration

