

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN103118\
 Data File : VN052085.D
 Acq On : 31 Oct 2018 20:19
 Operator : MD\SY
 Sample : J5202-12
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PL-02-103018-C

Quant Time: Nov 01 02:34:24 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N103018W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 31 02:45:22 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	289456	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	535758	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	503290	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	203569	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	245960	52.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.90%	
35) Dibromofluoromethane	7.59	113	205664	61.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	122.72%	
50) Toluene-d8	10.09	98	806082	60.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	120.34%	
62) 4-Bromofluorobenzene	12.40	95	274496	55.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.64%	
Target Compounds						
16) Acetone	3.82	43	11724	11.82	ug/l	90
20) Methylene Chloride	4.55	84	14127381	3739.46	ug/l	99
43) Isopropyl Acetate	8.17	43	164103	25.30	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN103118\
Data File : VN052085.D
Acq On : 31 Oct 2018 20:19
Operator : MD\SY
Sample : J5202-12
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PL-02-103018-C

Quant Time: Nov 01 02:34:24 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N103018W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 31 02:45:22 2018
Response via : Initial Calibration

