

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN103119\
 Data File : VN059029.D
 Acq On : 31 Oct 2019 11:20
 Operator : JC/SP
 Sample : PB124436TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB124436TB

Quant Time: Nov 01 00:44:37 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 19 00:16:20 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	410198	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	646602	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	572615	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	200186	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	276480	45.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.92%	
35) Dibromofluoromethane	7.57	113	207080	48.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.42%	
50) Toluene-d8	10.08	98	809944	49.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.16%	
62) 4-Bromofluorobenzene	12.41	95	251821	41.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.10%	
Target Compounds						
20) Methylene Chloride	4.52	84	8673	1.709	ug/l	97
43) Isopropyl Acetate	8.15	43	100870	8.720	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN103119\
Data File : VN059029.D
Acq On : 31 Oct 2019 11:20
Operator : JC/SP
Sample : PB124436TB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB124436TB

Quant Time: Nov 01 00:44:37 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 19 00:16:20 2019
Response via : Initial Calibration

