

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN103119\
 Data File : VN059031.D
 Acq On : 31 Oct 2019 12:04
 Operator : JC/SP
 Sample : K5143-09
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 NB-08-103019

Quant Time: Nov 01 01:06:18 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 19 00:16:20 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	409838	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	638687	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	583672	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	215309	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	275848	45.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.78%	
35) Dibromofluoromethane	7.57	113	205510	48.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.88%	
50) Toluene-d8	10.08	98	812735	49.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.72%	
62) 4-Bromofluorobenzene	12.41	95	265766	44.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.80%	
Target Compounds						
18) Methyl Acetate	4.31	43	9154	1.356	ug/l #	74
20) Methylene Chloride	4.53	84	522072	102.937	ug/l	97
43) Isopropyl Acetate	8.15	43	101750	8.905	ug/l #	93
95) Naphthalene	15.13	128	8688	2.898	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN103119\
Data File : VN059031.D
Acq On : 31 Oct 2019 12:04
Operator : JC/SP
Sample : K5143-09
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
NB-08-103019

Quant Time: Nov 01 01:06:18 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 19 00:16:20 2019
Response via : Initial Calibration

