

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN103119\
 Data File : VN059034.D
 Acq On : 31 Oct 2019 13:10
 Operator : JC/SP
 Sample : K5615-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RBR-29

Quant Time: Nov 01 02:13:21 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 19 00:16:20 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	398579	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	623120	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	568815	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	210373	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	272636	46.64	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.28%	
35) Dibromofluoromethane	7.57	113	200930	49.04	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.08%	
50) Toluene-d8	10.08	98	787665	49.53	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.06%	
62) 4-Bromofluorobenzene	12.41	95	252664	43.26	ug/l	0.00
Spiked Amount	50.000		Recovery	=	86.52%	
Target Compounds						
8) Diethyl Ether	3.39	74	5615	2.149	ug/l	94
16) Acetone	3.80	43	27569	9.544	ug/l	99
18) Methyl Acetate	4.31	43	7972	1.214	ug/l #	76
20) Methylene Chloride	4.53	84	288451	58.481	ug/l	98
43) Isopropyl Acetate	8.15	43	95337	8.552	ug/l #	92
95) Naphthalene	15.14	128	15372	3.444	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN103119\
Data File : VN059034.D
Acq On : 31 Oct 2019 13:10
Operator : JC/SP
Sample : K5615-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RBR-29

Quant Time: Nov 01 02:13:21 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 19 00:16:20 2019
Response via : Initial Calibration

