

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN103119\
 Data File : VN059035.D
 Acq On : 31 Oct 2019 13:32
 Operator : JC/SP
 Sample : K5615-04
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RBR-52

Quant Time: Nov 01 02:24:10 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 19 00:16:20 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	394503	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	618828	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	551513	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	196682	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	268366	46.38	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.76%	
35) Dibromofluoromethane	7.57	113	200841	49.36	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.72%	
50) Toluene-d8	10.09	98	780750	49.43	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.86%	
62) 4-Bromofluorobenzene	12.41	95	243567	41.99	ug/l	0.00
Spiked Amount	50.000		Recovery	=	83.98%	
Target Compounds						
8) Diethyl Ether	3.40	74	5848	2.261	ug/l	96
16) Acetone	3.80	43	31342	10.962	ug/l	98
18) Methyl Acetate	4.31	43	10216	1.572	ug/l #	75
20) Methylene Chloride	4.53	84	82835	16.967	ug/l	97
43) Isopropyl Acetate	8.15	43	89790	8.110	ug/l #	92
95) Naphthalene	15.13	128	17026	3.669	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN103119\
Data File : VN059035.D
Acq On : 31 Oct 2019 13:32
Operator : JC/SP
Sample : K5615-04
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RBR-52

Quant Time: Nov 01 02:24:10 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 19 00:16:20 2019
Response via : Initial Calibration

