

Data Path : Z:\VOASRV\HPCHEM1\MSVOA\_N\DATA\VN103119\  
 Data File : VN059038.D  
 Acq On : 31 Oct 2019 14:39  
 Operator : JC/SP  
 Sample : K5612-06  
 Misc : 5.00mL/MSVOA\_N/WATER  
 ALS Vial : 16 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 GL-STOCKPILE

Quant Time: Nov 01 02:36:45 2019  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_N\METHODS\82N101819W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Oct 19 00:16:20 2019  
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	378181	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	589781	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	533621	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	182116	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.01	65	257609	46.44	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	92.88%	
35) Dibromofluoromethane	7.57	113	194185	50.08	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	100.16%	
50) Toluene-d8	10.09	98	747199	49.64	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	99.28%	
62) 4-Bromofluorobenzene	12.41	95	229712	41.56	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	83.12%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.80	43	17069	6.228	ug/l	89
18) Methyl Acetate	4.30	43	11828	1.898	ug/l	95
20) Methylene Chloride	4.53	84	1151284	246.001	ug/l	98
43) Isopropyl Acetate	8.15	43	89141	8.448	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA\_N\DATA\VN103119\  
 Data File : VN059038.D  
 Acq On : 31 Oct 2019 14:39  
 Operator : JC/SP  
 Sample : K5612-06  
 Misc : 5.00mL/MSVOA\_N/WATER  
 ALS Vial : 16 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 GL-STOCKPILE

Quant Time: Nov 01 02:36:45 2019  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_N\METHODS\82N101819W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Oct 19 00:16:20 2019  
 Response via : Initial Calibration

