

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN103119\
 Data File : VN059039.D
 Acq On : 31 Oct 2019 15:01
 Operator : JC/SP
 Sample : K5138-06
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 AU-08-103019

Quant Time: Nov 01 02:41:25 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 19 00:16:20 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	416604	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	658154	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	610861	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	224195	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	287066	46.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.96%	
35) Dibromofluoromethane	7.57	113	213182	49.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.54%	
50) Toluene-d8	10.09	98	840447	50.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.06%	
62) 4-Bromofluorobenzene	12.41	95	271781	44.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.12%	
Target Compounds						
16) Acetone	3.80	43	16676	5.523	ug/l	96
18) Methyl Acetate	4.30	43	12345	1.798	ug/l	93
20) Methylene Chloride	4.53	84	274457	53.236	ug/l	98
43) Isopropyl Acetate	8.15	43	100947	8.573	ug/l #	92
95) Naphthalene	15.13	128	20070	3.719	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN103119\
Data File : VN059039.D
Acq On : 31 Oct 2019 15:01
Operator : JC/SP
Sample : K5138-06
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
AU-08-103019

Quant Time: Nov 01 02:41:25 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 19 00:16:20 2019
Response via : Initial Calibration

