


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN103119\
 Data File : VN059023.D
 Acq On : 31 Oct 2019 8:19
 Operator : JC/SP
 Sample : BFB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
 Title : SW846 8260
 Last Update : Sat Oct 19 00:16:20 2019



AutoFind: Scans 3305, 3306, 3307; Background Corrected with Scan 3291

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.8	16715	PASS
75	95	30	60	51.1	45485	PASS
95	95	100	100	100.0	88933	PASS
96	95	5	9	7.2	6360	PASS
173	174	0.00	2	1.1	773	PASS
174	95	50	100	76.7	68226	PASS
175	174	5	9	7.5	5126	PASS
176	174	95	101	95.3	65018	PASS
177	176	5	9	6.8	4391	PASS