

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN103122\
 Data File : VN075079.D
 Acq On : 31 Oct 2022 11:51
 Operator : JC\MD
 Sample : VN1031MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1031MBL01

Quant Time: Nov 01 06:12:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	139187	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	255817	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	221062	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	85649	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	102921	54.880	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	109.760%
35) Dibromofluoromethane	8.171	113	79356	52.841	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.680%
50) Toluene-d8	10.571	98	315039	54.070	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	108.140%
62) 4-Bromofluorobenzene	12.847	95	94742	46.289	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.580%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN103122\
Data File : VN075079.D
Acq On : 31 Oct 2022 11:51
Operator : JC\MD
Sample : VN1031MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1031MBL01

Quant Time: Nov 01 06:12:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

