

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN103122\
 Data File : VN075084.D
 Acq On : 31 Oct 2022 14:10
 Operator : JC\MD
 Sample : N5316-12
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TCLP-VOC-GRAB-3-PILE-2

Quant Time: Nov 01 06:16:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	102472	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	201589	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	186797	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	72417	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	83538	60.505	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 121.000%	
35) Dibromofluoromethane	8.177	113	64521	54.520	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 109.040%	
50) Toluene-d8	10.571	98	249120	54.258	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 108.520%	
62) 4-Bromofluorobenzene	12.847	95	80262	49.764	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 99.520%	
Target Compounds						
					Qvalue	
16) Acetone	4.442	43	10062	17.211	ug/l	97
20) Methylene Chloride	5.289	84	44605	34.220	ug/l	92
37) Ethyl Acetate	7.571	43	15517	7.723	ug/l	97
43) Isopropyl Acetate	8.694	43	68853	22.092	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN103122\
Data File : VN075084.D
Acq On : 31 Oct 2022 14:10
Operator : JC\MD
Sample : N5316-12
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TCLP-VOC-GRAB-3-PILE-2

Quant Time: Nov 01 06:16:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

