

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN103122\
 Data File : VN075086.D
 Acq On : 31 Oct 2022 14:59
 Operator : JC\MD
 Sample : PB148643TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB148643TB

Quant Time: Nov 01 06:17:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	108352	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	213115	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	204770	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	77630	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	86587	59.310	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	118.620%
35) Dibromofluoromethane	8.177	113	67041	53.586	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.180%
50) Toluene-d8	10.571	98	269994	55.624	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	111.240%
62) 4-Bromofluorobenzene	12.853	95	88605	51.965	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.940%
Target Compounds						
						Qvalue
16) Acetone	4.448	43	18688	30.231	ug/l	95
20) Methylene Chloride	5.289	84	5100	3.700	ug/l #	83
37) Ethyl Acetate	7.571	43	17331	8.159	ug/l	99
43) Isopropyl Acetate	8.688	43	231246	70.184	ug/l #	68

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN103122\
Data File : VN075086.D
Acq On : 31 Oct 2022 14:59
Operator : JC\MD
Sample : PB148643TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB148643TB

Quant Time: Nov 01 06:17:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

