

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN103122\
 Data File : VN075094.D
 Acq On : 31 Oct 2022 18:11
 Operator : JC\MD
 Sample : N5316-09
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 COMPOSITE-PILE-2

Quant Time: Nov 01 06:22:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	98882	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	196247	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	178572	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	68069	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	79063	59.342	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	118.680%
35) Dibromofluoromethane	8.177	113	59397	51.557	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.120%
50) Toluene-d8	10.571	98	245203	54.859	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	109.720%
62) 4-Bromofluorobenzene	12.853	95	75320	47.971	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.940%
Target Compounds						
						Qvalue
16) Acetone	4.442	43	4593	8.142	ug/l	95
20) Methylene Chloride	5.289	84	45868	36.466	ug/l	88
37) Ethyl Acetate	7.571	43	15904	8.131	ug/l	97
43) Isopropyl Acetate	8.694	43	120125	39.592	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN103122\
Data File : VN075094.D
Acq On : 31 Oct 2022 18:11
Operator : JC\MD
Sample : N5316-09
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
COMPOSITE-PILE-2

Quant Time: Nov 01 06:22:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

