

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN103122\
 Data File : VN075095.D
 Acq On : 31 Oct 2022 18:35
 Operator : JC\MD
 Sample : N5377-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 LIQUID-WC-20221028

Quant Time: Nov 01 06:22:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	92490	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	179338	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	149711	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	54067	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	76267	61.200	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	122.400%
35) Dibromofluoromethane	8.177	113	52462	49.831	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.660%
50) Toluene-d8	10.571	98	212746	52.085	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	104.160%
62) 4-Bromofluorobenzene	12.853	95	61451	42.828	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	85.660%
Target Compounds						
16) Acetone	4.459	43	129	0.244	ug/l #	43
43) Isopropyl Acetate	8.741	43	5208	1.878	ug/l #	53

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN103122\
Data File : VN075095.D
Acq On : 31 Oct 2022 18:35
Operator : JC\MD
Sample : N5377-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
LIQUID-WC-20221028

Quant Time: Nov 01 06:22:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

