

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN103123\
 Data File : VN079736.D
 Acq On : 31 Oct 2023 11:13
 Operator : JC\MD
 Sample : VN1031MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1031MBL01

Quant Time: Nov 01 01:16:21 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103023W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 31 07:23:24 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	263698	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	480183	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	411678	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	116058	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	224228	52.950	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.900%
35) Dibromofluoromethane	8.171	113	175588	51.175	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.340%
50) Toluene-d8	10.570	98	600715	46.964	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.920%
62) 4-Bromofluorobenzene	12.853	95	154380	38.327	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	76.660%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN103123\
Data File : VN079736.D
Acq On : 31 Oct 2023 11:13
Operator : JC\MD
Sample : VN1031MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1031MBL01

Quant Time: Nov 01 01:16:21 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103023W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 31 07:23:24 2023
Response via : Initial Calibration

