

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN103123\
 Data File : VN079737.D
 Acq On : 31 Oct 2023 12:13
 Operator : JC\MD
 Sample : VN1031WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1031WBL01

Quant Time: Nov 01 01:16:50 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103023W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 31 07:23:24 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	258079	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	469345	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	408308	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	104294	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	218729	52.776	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.560%
35) Dibromofluoromethane	8.171	113	171978	51.280	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.560%
50) Toluene-d8	10.571	98	593319	47.457	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.920%
62) 4-Bromofluorobenzene	12.853	95	147349	37.426	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	74.860%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN103123\
Data File : VN079737.D
Acq On : 31 Oct 2023 12:13
Operator : JC\MD
Sample : VN1031WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1031WBL01

Quant Time: Nov 01 01:16:50 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103023W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 31 07:23:24 2023
Response via : Initial Calibration

