

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN103123\
 Data File : VN079738.D
 Acq On : 31 Oct 2023 12:36
 Operator : JC\MD
 Sample : VN1031WBL02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1031WBL02

Quant Time: Nov 01 01:17:11 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103023W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 31 07:23:24 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	206469	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	407952	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	401741	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	110448	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	183884	55.459	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	110.920%
35) Dibromofluoromethane	8.171	113	163121	55.959	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	111.920%
50) Toluene-d8	10.570	98	506407	46.601	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.200%
62) 4-Bromofluorobenzene	12.853	95	156014	45.591	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.180%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN103123\
Data File : VN079738.D
Acq On : 31 Oct 2023 12:36
Operator : JC\MD
Sample : VN1031WBL02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1031WBL02

Quant Time: Nov 01 01:17:11 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103023W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 31 07:23:24 2023
Response via : Initial Calibration

