

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN103125\
 Data File : VN088193.D
 Acq On : 31 Oct 2025 16:46
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Nov 01 00:37:42 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102325W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 25 00:15:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	264276	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	598834	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.841	117	577940	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	282587	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.559	65	267536	58.543	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	117.080%
35) Dibromofluoromethane	8.147	113	201439	50.072	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.140%
50) Toluene-d8	10.541	98	768885	49.601	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.200%
62) 4-Bromofluorobenzene	12.829	95	315493	57.630	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	115.260%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN103125\
Data File : VN088193.D
Acq On : 31 Oct 2025 16:46
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Nov 01 00:37:42 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102325W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 25 00:15:22 2025
Response via : Initial Calibration

