

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN110118\
 Data File : VN052101.D
 Acq On : 1 Nov 2018 14:14
 Operator : MD\SY
 Sample : J5754-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 BEL-03-SUMP-53-ST

Quant Time: Nov 02 07:00:57 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\624N110118W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Nov 02 05:39:38 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.20	128	57370	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.59	114	328161	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	283980	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	8.03	65	133580	31.37	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	104.57%
60) 4-Bromofluorobenzene	12.40	95	124004	27.30	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	91.00%
63) Toluene-d8	10.09	98	404262	30.27	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	100.90%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN110118\
Data File : VN052101.D
Acq On : 1 Nov 2018 14:14
Operator : MD\SY
Sample : J5754-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
BEL-03-SUMP-53-ST

Quant Time: Nov 02 07:00:57 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\624N110118W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Nov 02 05:39:38 2018
Response via : Initial Calibration

