

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110118\
 Data File : VN052103.D
 Acq On : 1 Nov 2018 15:14
 Operator : MD\SY
 Sample : VN1101WBL02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1101WBL02

Quant Time: Nov 02 07:13:08 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N103018W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 31 02:45:22 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	305945	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	571612	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	530385	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	214933	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	257892	52.53	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.06%	
35) Dibromofluoromethane	7.59	113	214743	60.05	ug/l	0.00
Spiked Amount	50.000		Recovery	=	120.10%	
50) Toluene-d8	10.09	98	842816	58.97	ug/l	0.00
Spiked Amount	50.000		Recovery	=	117.94%	
62) 4-Bromofluorobenzene	12.40	95	282451	53.83	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.66%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110118\
Data File : VN052103.D
Acq On : 1 Nov 2018 15:14
Operator : MD\SY
Sample : VN1101WBL02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
VN1101WBL02

Quant Time: Nov 02 07:13:08 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N103018W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 31 02:45:22 2018
Response via : Initial Calibration

