

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110118\
 Data File : VN052109.D
 Acq On : 1 Nov 2018 18:02
 Operator : MD\SY
 Sample : J5200-18
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 OR-03-103118-A

Quant Time: Nov 02 07:38:51 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N103018W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 31 02:45:22 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	291070	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	547778	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	527049	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	218479	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	251951	53.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.88%	
35) Dibromofluoromethane	7.59	113	207206	60.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	120.94%	
50) Toluene-d8	10.09	98	819922	59.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	119.72%	
62) 4-Bromofluorobenzene	12.40	95	288281	57.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.66%	
Target Compounds						
16) Acetone	3.82	43	22171	22.238	ug/l	100
20) Methylene Chloride	4.56	84	289433	76.187	ug/l	95
43) Isopropyl Acetate	8.17	43	176592	26.632	ug/l	98
68) m/p-Xylenes	11.62	106	13168	1.858	ug/l	95
69) o-Xylene	11.95	106	10107	1.442	ug/l	96
80) 1,3,5-Trimethylbenzene	12.73	105	15361	1.010	ug/l	100
84) 1,2,4-Trimethylbenzene	13.04	105	63253	4.180	ug/l	98
95) Naphthalene	15.13	128	35176	5.619	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110118\
Data File : VN052109.D
Acq On : 1 Nov 2018 18:02
Operator : MD\SY
Sample : J5200-18
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
OR-03-103118-A

Quant Time: Nov 02 07:38:51 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N103018W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 31 02:45:22 2018
Response via : Initial Calibration

