

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN110118\
 Data File : VN052102.D
 Acq On : 1 Nov 2018 14:41
 Operator : MD\SY
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 28

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 02 07:10:49 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N103018W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 31 02:45:22 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	134	0.00
2 T	Dichlorodifluoromethane	0.608	0.606	0.3	129	0.00
3 P	Chloromethane	0.766	0.692	9.7	120	0.00
4 C	Vinyl Chloride	0.824	0.700	15.0#	112	0.00
5 T	Bromomethane	0.618	0.472	23.6	108	0.00
6 T	Chloroethane	0.532	0.435	18.2	110	0.00
7 T	Trichlorofluoromethane	0.942	0.929	1.4	131	0.00
8 T	Diethyl Ether	0.351	0.331	5.7	130	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.555	0.550	0.9	132	0.00
10 T	Methyl Iodide	0.728	0.743	-2.1	135	0.00
11 T	Tert butyl alcohol	0.040	0.039	2.5	143	0.00
12 CM	1,1-Dichloroethene	0.525	0.507	3.4#	128	0.00
13 T	Acrolein	0.036	0.031	13.9	108	0.00
14 T	Allyl chloride	1.050	0.930	11.4	115	0.00
15 T	Acrylonitrile	0.230	0.206	10.4	121	0.00
16 T	Acetone	0.171	0.160	6.4	132	0.00
17 T	Carbon Disulfide	1.653	1.473	10.9	120	0.00
18 T	Methyl Acetate	0.545	0.470	13.8	128	0.00
19 T	Methyl tert-butyl Ether	1.723	1.557	9.6	122	0.00
20 T	Methylene Chloride	0.653	0.621	4.9	127	0.00
21 T	trans-1,2-Dichloroethene	0.598	0.577	3.5	126	0.00
22 T	Diisopropyl ether	2.269	1.973	13.0	115	0.00
23 T	Vinyl Acetate	1.572	1.414	10.1	116	0.00
24 P	1,1-Dichloroethane	1.269	1.102	13.2	116	0.00
25 T	2-Butanone	0.269	0.252	6.3	126	0.00
26 T	2,2-Dichloropropane	1.008	0.842	16.5	109	0.00
27 T	cis-1,2-Dichloroethene	0.714	0.680	4.8	125	0.00
28 T	Bromochloromethane	0.629	0.504	19.9	113	0.00
29 T	Tetrahydrofuran	0.179	0.167	6.7	123	0.00
30 C	Chloroform	1.266	1.132	10.6#	119	0.00
31 T	Cyclohexane	1.507	1.028	31.8#	113	0.00
32 T	1,1,1-Trichloroethane	1.018	0.984	3.3	126	0.00
33 S	1,2-Dichloroethane-d4	0.802	0.654	18.5	112	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	126	0.00
35 S	Dibromofluoromethane	0.313	0.312	0.3	131	0.00
36 T	1,1-Dichloropropene	0.508	0.475	6.5	118	0.00
37 T	Ethyl Acetate	0.340	0.310	8.8	126	0.00
38 T	Carbon Tetrachloride	0.438	0.473	-8.0	132	0.00
39 T	Methylcyclohexane	0.603	0.581	3.6	117	0.00
40 TM	Benzene	1.471	1.409	4.2	119	0.00
41 T	Methacrylonitrile	0.145	0.161	-11.0	109	0.00
42 TM	1,2-Dichloroethane	0.543	0.465	14.4	112	0.00
43 T	Isopropyl Acetate	0.605	0.560	7.4	115	0.00
44 TM	Trichloroethene	0.329	0.349	-6.1	131	0.00
45 C	1,2-Dichloropropane	0.419	0.374	10.7#	115	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN110118\
 Data File : VN052102.D
 Acq On : 1 Nov 2018 14:41
 Operator : MD\SY
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 28

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 02 07:10:49 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N103018W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 31 02:45:22 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.234	0.230	1.7	120	0.00
47 T	Bromodichloromethane	0.488	0.477	2.3	117	0.00
48 T	Methyl methacrylate	0.307	0.285	7.2	116	0.00
49 T	1,4-Dioxane	0.002	0.003	-50.0#	156#	0.00
50 S	Toluene-d8	1.250	1.186	5.1	123	0.00
51 T	4-Methyl-2-Pentanone	0.329	0.319	3.0	120	0.00
52 CM	Toluene	0.888	0.871	1.9#	120	0.00
53 T	t-1,3-Dichloropropene	0.516	0.480	7.0	113	0.00
54 T	cis-1,3-Dichloropropene	0.573	0.553	3.5	115	0.00
55 T	1,1,2-Trichloroethane	0.316	0.316	0.0	123	0.00
56 T	Ethyl methacrylate	0.452	0.429	5.1	116	0.00
57 T	1,3-Dichloropropane	0.582	0.536	7.9	115	0.00
58 T	2-Chloroethyl Vinyl ether	0.241	0.228	5.4	116	0.00
59 T	2-Hexanone	0.234	0.215	8.1	116	0.00
60 T	Dibromochloromethane	0.311	0.343	-10.3	131	0.00
61 T	1,2-Dibromoethane	0.314	0.321	-2.2	125	0.00
62 S	4-Bromofluorobenzene	0.459	0.420	8.5	117	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	122	0.00
64 T	Tetrachloroethene	0.303	0.324	-6.9	132	0.00
65 PM	Chlorobenzene	1.053	1.063	-0.9	122	0.00
66 T	1,1,1,2-Tetrachloroethane	0.331	0.378	-14.2	131	0.00
67 C	Ethyl Benzene	1.878	1.862	0.9#	116	0.00
68 T	m/p-Xylenes	0.672	0.716	-6.5	121	0.00
69 T	o-Xylene	0.665	0.698	-5.0	119	0.00
70 T	Styrene	1.076	1.128	-4.8	118	0.00
71 P	Bromoform	0.187	0.223	-19.3	134	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	121	0.00
73 T	Isopropylbenzene	4.276	4.171	2.5	117	0.00
74 T	N-amyl acetate	1.492	1.362	8.7	116	0.00
75 P	1,1,2,2-Tetrachloroethane	1.025	0.963	6.0	118	0.00
76 T	1,2,3-Trichloropropane	0.915	0.888	3.0	115	0.00
77 T	Bromobenzene	0.895	0.975	-8.9	131	0.00
78 T	n-propylbenzene	5.181	4.846	6.5	112	0.00
79 T	2-Chlorotoluene	3.020	2.812	6.9	112	0.00
80 T	1,3,5-Trimethylbenzene	3.479	3.419	1.7	117	0.00
81 T	trans-1,4-Dichloro-2-butene	0.276	0.258	6.5	109	0.00
82 T	4-Chlorotoluene	3.080	2.893	6.1	111	0.00
83 T	tert-Butylbenzene	2.923	2.990	-2.3	121	0.00
84 T	1,2,4-Trimethylbenzene	3.463	3.415	1.4	115	0.00
85 T	sec-Butylbenzene	4.175	4.036	3.3	114	0.00
86 T	p-Isopropyltoluene	3.400	3.433	-1.0	117	0.00
87 T	1,3-Dichlorobenzene	1.631	1.718	-5.3	127	0.00
88 T	1,4-Dichlorobenzene	1.635	1.696	-3.7	125	0.00
89 T	n-Butylbenzene	3.243	2.945	9.2	107	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN110118\
Data File : VN052102.D
Acq On : 1 Nov 2018 14:41
Operator : MD\SY
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 28

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Nov 02 07:10:49 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N103018W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 31 02:45:22 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.583	0.582	0.2	117	0.00
91 T	1,2-Dichlorobenzene	1.534	1.640	-6.9	126	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.127	0.140	-10.2	117	0.00
93 T	1,2,4-Trichlorobenzene	0.571	0.693	-21.4	143	0.00
94 T	Hexachlorobutadiene	0.320	0.337	-5.3	134	0.00
95 T	Naphthalene	1.433	1.673	-16.7	133	0.00
96 T	1,2,3-Trichlorobenzene	0.493	0.595	-20.7	143	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6