

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN110119\
 Data File : VN059063.D
 Acq On : 1 Nov 2019 15:37
 Operator : JC/SP
 Sample : K5626-10DL 5X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SS003RW114R-191030DL

Quant Time: Nov 02 05:28:29 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 02 05:43:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	427393	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	656187	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	568818	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	212441	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	290942	51.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.80%	
35) Dibromofluoromethane	7.57	113	215284	50.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.72%	
50) Toluene-d8	10.08	98	854030	51.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.64%	
62) 4-Bromofluorobenzene	12.40	95	276874	46.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.46%	
Target Compounds						
65) Chlorobenzene	11.43	112	74632	6.393	ug/l	100
87) 1,3-Dichlorobenzene	13.29	146	15150	2.182	ug/l	99
88) 1,4-Dichlorobenzene	13.37	146	40785	5.938	ug/l	94
91) 1,2-Dichlorobenzene	13.66	146	215483	31.190	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN110119\
Data File : VN059063.D
Acq On : 1 Nov 2019 15:37
Operator : JC/SP
Sample : K5626-10DL 5X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SS003RW114R-191030DL

Quant Time: Nov 02 05:28:29 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N110119W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 02 05:43:23 2019
Response via : Initial Calibration

