

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN110119\
 Data File : VN059064.D
 Acq On : 1 Nov 2019 15:58
 Operator : JC/SP
 Sample : PB124481TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB124481TB

Quant Time: Nov 02 05:31:56 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 02 05:43:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	438138	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	678123	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	577908	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	203035	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	287391	49.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.06%	
35) Dibromofluoromethane	7.57	113	214899	49.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.24%	
50) Toluene-d8	10.08	98	823115	48.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.66%	
62) 4-Bromofluorobenzene	12.40	95	255309	41.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.40%	
Target Compounds						
20) Methylene Chloride	4.52	84	8401	1.634	ug/l	Qvalue 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN110119\
Data File : VN059064.D
Acq On : 1 Nov 2019 15:58
Operator : JC/SP
Sample : PB124481TB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB124481TB

Quant Time: Nov 02 05:31:56 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N110119W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 02 05:43:23 2019
Response via : Initial Calibration

