

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110121\
 Data File : VN069318.D
 Acq On : 1 Nov 2021 13:57
 Operator : JC/MD
 Sample : VN1101WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1101WBL01

Quant Time: Nov 01 17:36:54 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102821W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 28 16:09:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	324875	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	582948	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	570696	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	218216	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	225414	48.066	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	96.140%
35) Dibromofluoromethane	8.024	113	167927	46.211	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	92.420%
50) Toluene-d8	10.443	98	725358	49.923	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	99.840%
62) 4-Bromofluorobenzene	12.734	95	269061	49.074	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	98.140%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110121\
Data File : VN069318.D
Acq On : 1 Nov 2021 13:57
Operator : JC/MD
Sample : VN1101WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1101WBL01

Quant Time: Nov 01 17:36:54 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102821W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 28 16:09:15 2021
Response via : Initial Calibration

