

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110121\
 Data File : VN069332.D
 Acq On : 1 Nov 2021 20:24
 Operator : JC/MD
 Sample : M4407-03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TRIP-BLANK

Quant Time: Nov 02 05:51:24 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102821W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 28 16:09:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	285623	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	522499	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	524943	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	213816	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	200363	48.595	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.200%
35) Dibromofluoromethane	8.024	113	150738	46.280	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	92.560%
50) Toluene-d8	10.443	98	653486	50.180	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	100.360%
62) 4-Bromofluorobenzene	12.733	95	249734	50.819	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	101.640%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110121\
Data File : VN069332.D
Acq On : 1 Nov 2021 20:24
Operator : JC/MD
Sample : M4407-03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TRIP-BLANK

Quant Time: Nov 02 05:51:24 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102821W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 28 16:09:15 2021
Response via : Initial Calibration

