

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110121\
 Data File : VN069333.D
 Acq On : 1 Nov 2021 20:49
 Operator : JC/MD
 Sample : M4427-09
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TRIP-BLANK

Quant Time: Nov 02 05:51:34 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102821W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 28 16:09:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	285845	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	513073	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	515513	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	213965	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	201387	48.806	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.620%
35) Dibromofluoromethane	8.024	113	151074	47.235	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	94.480%
50) Toluene-d8	10.443	98	642727	50.261	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	100.520%
62) 4-Bromofluorobenzene	12.734	95	248640	51.526	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	103.060%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110121\
Data File : VN069333.D
Acq On : 1 Nov 2021 20:49
Operator : JC/MD
Sample : M4427-09
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TRIP-BLANK

Quant Time: Nov 02 05:51:34 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102821W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 28 16:09:15 2021
Response via : Initial Calibration

