

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110122\
 Data File : VN075102.D
 Acq On : 01 Nov 2022 12:33
 Operator : JC\MD
 Sample : VN1101WBSD01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1101WBSD01

Quant Time: Nov 02 07:10:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	88070	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	165188	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	150176	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	72181	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	65837	55.482	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	110.960%
35) Dibromofluoromethane	8.177	113	50604	52.183	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.360%
50) Toluene-d8	10.571	98	194105	51.592	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.180%
62) 4-Bromofluorobenzene	12.853	95	67619	51.163	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.320%
Target Compounds						Qvalue
2) Dichlorodifluoromethane	2.136	85	22258	19.428	ug/l	92
3) Chloromethane	2.377	50	24069	19.778	ug/l	94
4) Vinyl Chloride	2.524	62	32789	18.887	ug/l	92
5) Bromomethane	2.936	94	27818	19.411	ug/l	100
6) Chloroethane	3.112	64	24512	19.690	ug/l	99
7) Trichlorofluoromethane	3.500	101	34759	19.600	ug/l	95
8) Diethyl Ether	3.977	74	13420	20.746	ug/l	97
9) 1,1,2-Trichlorotrifluo...	4.383	101	20378	20.807	ug/l	98
10) Methyl Iodide	4.600	142	28012	19.212	ug/l	99
11) Tert butyl alcohol	5.542	59	27388	98.971	ug/l	98
12) 1,1-Dichloroethene	4.347	96	18084	19.371	ug/l	93
13) Acrolein	4.195	56	25040	110.626	ug/l	100
14) Allyl chloride	5.030	41	27278	20.006	ug/l	97
15) Acrylonitrile	5.730	53	61874	105.845	ug/l	98
16) Acetone	4.453	43	54358	108.185	ug/l	93
17) Carbon Disulfide	4.718	76	42860	18.256	ug/l #	95
18) Methyl Acetate	5.036	43	37017	22.554	ug/l	96
19) Methyl tert-butyl Ether	5.806	73	71123	20.498	ug/l	97
20) Methylene Chloride	5.283	84	23148	20.662	ug/l	91
21) trans-1,2-Dichloroethene	5.800	96	19510	19.087	ug/l	92
22) Diisopropyl ether	6.689	45	63180	20.997	ug/l	96
23) Vinyl Acetate	6.612	43	230313	91.662	ug/l	98
24) 1,1-Dichloroethane	6.583	63	38822	20.400	ug/l	98
25) 2-Butanone	7.494	43	83121	105.176	ug/l	98
26) 2,2-Dichloropropane	7.500	77	32575	18.607	ug/l	99
27) cis-1,2-Dichloroethene	7.500	96	24824	20.336	ug/l	98
28) Bromochloromethane	7.824	49	17786	22.171	ug/l	94
29) Tetrahydrofuran	7.847	42	53562	104.882	ug/l	98
30) Chloroform	7.977	83	41691	20.178	ug/l	97
31) Cyclohexane	8.265	56	34770	19.701	ug/l	96
32) 1,1,1-Trichloroethane	8.177	97	36928	19.681	ug/l	96
36) 1,1-Dichloropropene	8.377	75	28692	19.111	ug/l	99
37) Ethyl Acetate	7.565	43	33080	20.091	ug/l	98
38) Carbon Tetrachloride	8.371	117	31299	18.952	ug/l	99
39) Methylcyclohexane	9.600	83	37147	18.742	ug/l	96
40) Benzene	8.612	78	90081	19.596	ug/l	96

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110122\
 Data File : VN075102.D
 Acq On : 01 Nov 2022 12:33
 Operator : JC\MD
 Sample : VN1101WBSD01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1101WBSD01

Quant Time: Nov 02 07:10:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.794	41	17226	21.838	ug/l	96
42) 1,2-Dichloroethane	8.677	62	34480	20.617	ug/l	99
43) Isopropyl Acetate	8.694	43	53467	20.935	ug/l	98
44) Trichloroethene	9.359	130	22240	19.284	ug/l	97
45) 1,2-Dichloropropane	9.624	63	23310	21.308	ug/l	95
46) Dibromomethane	9.712	93	17459	20.485	ug/l	96
47) Bromodichloromethane	9.894	83	33307	20.180	ug/l	96
48) Methyl methacrylate	9.682	41	24601	20.999	ug/l	97
49) 1,4-Dioxane	9.694	88	12750	415.616	ug/l	97
51) 4-Methyl-2-Pentanone	10.447	43	173694	105.394	ug/l	99
52) Toluene	10.635	92	59892	19.491	ug/l	96
53) t-1,3-Dichloropropene	10.841	75	34234	19.651	ug/l	94
54) cis-1,3-Dichloropropene	10.312	75	36929	19.986	ug/l	99
55) 1,1,2-Trichloroethane	11.024	97	24467	20.081	ug/l	92
56) Ethyl methacrylate	10.876	69	36658	19.724	ug/l	95
57) 1,3-Dichloropropane	11.165	76	42318	20.856	ug/l	99
58) 2-Chloroethyl Vinyl ether	10.165	63	77720	108.169	ug/l	98
59) 2-Hexanone	11.200	43	131010	104.741	ug/l	98
60) Dibromochloromethane	11.359	129	25537	19.866	ug/l	100
61) 1,2-Dibromoethane	11.471	107	25053	20.064	ug/l	99
64) Tetrachloroethene	11.106	164	19548	21.232	ug/l	94
65) Chlorobenzene	11.894	112	64841	20.228	ug/l	98
66) 1,1,1,2-Tetrachloroethane	11.965	131	23568	19.820	ug/l	99
67) Ethyl Benzene	11.965	91	113537	19.687	ug/l	99
68) m/p-Xylenes	12.070	106	89918	38.857	ug/l	94
69) o-Xylene	12.400	106	44725	19.543	ug/l	98
70) Styrene	12.412	104	71204	19.513	ug/l	98
71) Bromoform	12.582	173	17096	19.203	ug/l #	97
73) Isopropylbenzene	12.700	105	112716	19.451	ug/l	100
74) N-amyl acetate	12.494	43	44376	20.705	ug/l	97
75) 1,1,2,2-Tetrachloroethane	12.935	83	40181	21.887	ug/l	99
76) 1,2,3-Trichloropropane	12.994	75	48936	31.930	ug/l	72
77) Bromobenzene	12.982	156	25335	19.856	ug/l	96
78) n-propylbenzene	13.035	91	129431	19.894	ug/l	98
79) 2-Chlorotoluene	13.129	91	80674	20.292	ug/l	99
80) 1,3,5-Trimethylbenzene	13.176	105	98257	20.188	ug/l	98
81) trans-1,4-Dichloro-2-b...	12.741	75	10820	18.123	ug/l	94
82) 4-Chlorotoluene	13.129	91	80674	20.292	ug/l	99
83) tert-Butylbenzene	13.441	119	85187	19.697	ug/l	97
84) 1,2,4-Trimethylbenzene	13.482	105	98368	19.975	ug/l	98
85) sec-Butylbenzene	13.617	105	118985	19.667	ug/l	97
86) p-Isopropyltoluene	13.729	119	97878	19.589	ug/l	99
87) 1,3-Dichlorobenzene	13.735	146	48819	19.541	ug/l	97
88) 1,4-Dichlorobenzene	13.812	146	49656	20.084	ug/l	99
89) n-Butylbenzene	14.059	91	80095	19.906	ug/l	99
90) Hexachloroethane	14.335	117	17202	18.777	ug/l	96
91) 1,2-Dichlorobenzene	14.106	146	50069	20.160	ug/l	97
92) 1,2-Dibromo-3-Chloropr...	14.723	75	8227	21.048	ug/l	86
93) 1,2,4-Trichlorobenzene	15.394	180	20418	19.105	ug/l	98
94) Hexachlorobutadiene	15.506	225	10014	19.220	ug/l	98
95) Naphthalene	15.641	128	79298	19.294	ug/l	99
96) 1,2,3-Trichlorobenzene	15.841	180	21726	20.459	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110122\
Data File : VN075102.D
Acq On : 01 Nov 2022 12:33
Operator : JC\MD
Sample : VN1101WBSD01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1101WBSD01


Quant Time: Nov 02 07:10:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Nov 02 07:10:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 8.235 min Scan# 10

Delta R.T. 0.005 min

Lab File: VN075102.D

Acq: 01 Nov 2022 12:33

Instrument :

MSVOA_N

ClientSampleId :





VN1101WBSD01

Tgt Ion:168 Resp: 88070

Ion Ratio Lower Upper

168 100

99 60.0 51.6 77.4



#2

Dichlorodifluoromethane

Concen: 19.428 ug/l

RT: 2.136 min Scan# 14

Delta R.T. -0.000 min

Lab File: VN075102.D




Acq: 01 Nov 2022 12:33


Tgt Ion: 85 Resp: 22258

Ion Ratio Lower Upper

85 100

87 28.7 16.7 50.0









#3
Chloromethane
Concen: 19.778 ug/l
RT: 2.377 min Scan# 51
Delta R.T. 0.006 min
Lab File: VN075102.D
Acq: 01 Nov 2022 12:33




Instrument :
MSVOA_N
ClientSampleId :
VN1101WBSD01


Tgt Ion: 50 Resp: 24069
Ion Ratio Lower Upper
50 100
52 35.4 25.8 38.6



#4
Vinyl Chloride
Concen: 18.887 ug/l
RT: 2.524 min Scan# 80
Delta R.T. 0.000 min
Lab File: VN075102.D
Acq: 01 Nov 2022 12:33

Tgt Ion: 62 Resp: 32789
Ion Ratio Lower Upper
62 100
64 34.4 24.0 36.0









#5
Bromomethane
Concen: 19.411 ug/l
RT: 2.936 min Scan# 11
Delta R.T. 0.018 min
Lab File: VN075102.D
Acq: 01 Nov 2022 12:33




Instrument :
MSVOA_N
ClientSampleId :
VN1101WBSD01


Tgt Ion: 94 Resp: 27818
Ion Ratio Lower Upper
94 100
96 89.5 71.8 107.6



#6
Chloroethane
Concen: 19.690 ug/l
RT: 3.112 min Scan# 180
Delta R.T. 0.017 min
Lab File: VN075102.D
Acq: 01 Nov 2022 12:33

Tgt Ion: 64 Resp: 24512
Ion Ratio Lower Upper
64 100
66 30.6 23.8 35.8









#7
 Trichlorofluoromethane
 Concen: 19.600 ug/l
 RT: 3.500 min Scan# 244
 Delta R.T. 0.011 min
 Lab File: VN075102.D
 Acq: 01 Nov 2022 12:33




Instrument :
 MSVOA_N
 ClientSampleId :
 VN1101WBSD01


Tgt Ion:101 Resp: 34759
 Ion Ratio Lower Upper
 101 100
 103 66.7 50.5 75.7



#8
 Diethyl Ether
 Concen: 20.746 ug/l
 RT: 3.977 min Scan# 327
 Delta R.T. 0.006 min
 Lab File: VN075102.D
 Acq: 01 Nov 2022 12:33

Tgt Ion: 74 Resp: 13420
 Ion Ratio Lower Upper
 74 100
 45 80.9 39.3 117.9









#9
1,1,2-Trichlorotrifluoroethane
Concen: 20.807 ug/l
RT: 4.383 min Scan# 396
Delta R.T. -0.000 min
Lab File: VN075102.D
Acq: 01 Nov 2022 12:33




Instrument :
MSVOA_N
ClientSampleId :
VN1101WBSD01


Tgt Ion:101 Resp: 20378
Ion Ratio Lower Upper
101 100
85 46.4 37.6 56.4
151 71.0 58.3 87.5



#10
Methyl Iodide
Concen: 19.212 ug/l
RT: 4.600 min Scan# 433
Delta R.T. 0.005 min
Lab File: VN075102.D
Acq: 01 Nov 2022 12:33

Tgt Ion:142 Resp: 28012
Ion Ratio Lower Upper
142 100
127 47.1 38.0 57.0
141 14.7 12.2 18.4









#11
Tert butyl alcohol
Concen: 98.971 ug/l
RT: 5.542 min Scan# 591
Delta R.T. -0.000 min
Lab File: VN075102.D
Acq: 01 Nov 2022 12:33




Instrument :
MSVOA_N
ClientSampleId :
VN1101WBSD01


Tgt Ion: 59 Resp: 27388
Ion Ratio Lower Upper
59 100
57 11.2 8.5 12.7



#12
1,1-Dichloroethene
Concen: 19.371 ug/l
RT: 4.347 min Scan# 390
Delta R.T. -0.001 min
Lab File: VN075102.D
Acq: 01 Nov 2022 12:33

Tgt Ion: 96 Resp: 18084
Ion Ratio Lower Upper
96 100
61 150.8 126.2 189.4
98 54.8 51.5 77.3









#13
Acrolein
Concen: 110.626 ug/l
RT: 4.195 min Scan# 364
Delta R.T. 0.006 min
Lab File: VN075102.D
Acq: 01 Nov 2022 12:33




Instrument :
MSVOA_N
ClientSampleId :
VN1101WBSD01


Tgt Ion: 56 Resp: 25040
Ion Ratio Lower Upper
56 100
55 68.9 55.2 82.8



#14
Allyl chloride
Concen: 20.006 ug/l
RT: 5.030 min Scan# 506
Delta R.T. 0.006 min
Lab File: VN075102.D
Acq: 01 Nov 2022 12:33

Tgt Ion: 41 Resp: 27278
Ion Ratio Lower Upper
41 100
39 74.9 61.0 91.4
76 38.7 33.4 50.0









#15
Acrylonitrile
Concen: 105.845 ug/l
RT: 5.730 min Scan# 61
Delta R.T. -0.000 min
Lab File: VN075102.D
Acq: 01 Nov 2022 12:33




Instrument :
MSVOA_N
ClientSampleId :
VN1101WBSD01


Tgt Ion:	53	Resp:	61874
Ion	Ratio	Lower	Upper
53	100		
52	85.5	66.4	99.6
51	36.9	29.6	44.4



#16
Acetone
Concen: 108.185 ug/l
RT: 4.453 min Scan# 408
Delta R.T. 0.011 min
Lab File: VN075102.D
Acq: 01 Nov 2022 12:33

Tgt Ion:	43	Resp:	54358
Ion	Ratio	Lower	Upper
43	100		
58	28.3	25.5	38.3









#17
Carbon Disulfide
Concen: 18.256 ug/l
RT: 4.718 min Scan# 41
Delta R.T. 0.000 min
Lab File: VN075102.D
Acq: 01 Nov 2022 12:33




Instrument :
MSVOA_N
ClientSampleId :
VN1101WBSD01


Tgt Ion: 76 Resp: 42860
Ion Ratio Lower Upper
76 100
78 7.4 7.4 11.2#



#18
Methyl Acetate
Concen: 22.554 ug/l
RT: 5.036 min Scan# 507
Delta R.T. -0.006 min
Lab File: VN075102.D
Acq: 01 Nov 2022 12:33

Tgt Ion: 43 Resp: 37017
Ion Ratio Lower Upper
43 100
74 29.7 22.0 33.0









#19
Methyl tert-butyl Ether
Concen: 20.498 ug/l
RT: 5.806 min Scan# 611
Delta R.T. 0.000 min
Lab File: VN075102.D
Acq: 01 Nov 2022 12:33




Instrument :
MSVOA_N
ClientSampleId :
VN1101WBSD01


Tgt Ion: 73 Resp: 71123
Ion Ratio Lower Upper
73 100
57 19.0 16.3 24.5



#20
Methylene Chloride
Concen: 20.662 ug/l
RT: 5.283 min Scan# 549
Delta R.T. -0.000 min
Lab File: VN075102.D
Acq: 01 Nov 2022 12:33

Tgt Ion: 84 Resp: 23148
Ion Ratio Lower Upper
84 100
49 105.5 94.9 142.3
51 34.7 26.8 40.2
86 62.1 53.7 80.5









#21
trans-1,2-Dichloroethene
Concen: 19.087 ug/l
RT: 5.800 min Scan# 61
Delta R.T. 0.000 min
Lab File: VN075102.D
Acq: 01 Nov 2022 12:33




Instrument :
MSVOA_N
ClientSampleId :
VN1101WBSD01


Tgt Ion: 96 Resp: 19510
Ion Ratio Lower Upper
96 100
61 112.8 99.9 149.9
98 62.6 51.5 77.3



#22
Diisopropyl ether
Concen: 20.997 ug/l
RT: 6.689 min Scan# 788
Delta R.T. 0.006 min
Lab File: VN075102.D
Acq: 01 Nov 2022 12:33

Tgt Ion: 45 Resp: 63180
Ion Ratio Lower Upper
45 100
43 55.7 42.8 64.2
87 31.6 28.4 42.6
59 12.2 10.4 15.6









#23
 Vinyl Acetate
 Concen: 91.662 ug/l
 RT: 6.612 min Scan# 775
 Delta R.T. 0.000 min
 Lab File: VN075102.D
 Acq: 01 Nov 2022 12:33




Instrument :
 MSVOA_N
 ClientSampleId :
 VN1101WBSD01


Tgt Ion: 43 Resp: 230313
 Ion Ratio Lower Upper
 43 100
 86 12.5 10.7 16.1



#24
 1,1-Dichloroethane
 Concen: 20.400 ug/l
 RT: 6.583 min Scan# 770
 Delta R.T. 0.006 min
 Lab File: VN075102.D
 Acq: 01 Nov 2022 12:33

Tgt Ion: 63 Resp: 38822
 Ion Ratio Lower Upper
 63 100
 98 8.3 3.6 10.8
 100 4.4 2.4 7.2









#25
2-Butanone
Concen: 105.176 ug/l
RT: 7.494 min Scan# 912
Delta R.T. 0.000 min
Lab File: VN075102.D
Acq: 01 Nov 2022 12:33




Instrument :
MSVOA_N
ClientSampleId :
VN1101WBSD01


Tgt Ion: 43 Resp: 83121
Ion Ratio Lower Upper
43 100
72 30.9 25.5 38.3



#26
2,2-Dichloropropane
Concen: 18.607 ug/l
RT: 7.500 min Scan# 926
Delta R.T. 0.000 min
Lab File: VN075102.D
Acq: 01 Nov 2022 12:33

Tgt Ion: 77 Resp: 32575
Ion Ratio Lower Upper
77 100
97 22.7 11.2 33.6





#27

cis-1,2-Dichloroethene

Concen: 20.336 ug/l

RT: 7.500 min Scan# 91

Delta R.T. 0.011 min

Lab File: VN075102.D

Acq: 01 Nov 2022 12:33

Instrument :

MSVOA_N

ClientSampleId :

VN1101WBSD01





Tgt Ion: 96 Resp: 24824

Ion Ratio Lower Upper

96 100

61 138.4 0.0 278.8

98 61.4 0.0 128.4



#28

Bromochloromethane

Concen: 22.171 ug/l

RT: 7.824 min Scan# 981

Delta R.T. 0.006 min

Lab File: VN075102.D

Acq: 01 Nov 2022 12:33




Tgt Ion: 49 Resp: 17786


Ion Ratio Lower Upper

49 100

129 0.7 0.0 4.6

130 77.2 66.2 99.4





#29

Tetrahydrofuran

Concen: 104.882 ug/l

RT: 7.847 min Scan# 91

Delta R.T. 0.000 min

Lab File: VN075102.D

Acq: 01 Nov 2022 12:33

Instrument :

MSVOA_N

ClientSampleId :

VN1101WBSD01





Tgt Ion: 42 Resp: 53562

Ion Ratio Lower Upper

42 100

72 51.7 43.0 64.6

71 50.0 39.9 59.9



#30

Chloroform

Concen: 20.178 ug/l

RT: 7.977 min Scan# 1007

Delta R.T. 0.006 min

Lab File: VN075102.D




Acq: 01 Nov 2022 12:33


Tgt Ion: 83 Resp: 41691

Ion Ratio Lower Upper

83 100

85 61.5 51.4 77.0









#31
Cyclohexane
Concen: 19.701 ug/l
RT: 8.265 min Scan# 1055
Delta R.T. 0.006 min
Lab File: VN075102.D
Acq: 01 Nov 2022 12:33




Instrument :
MSVOA_N
ClientSampleId :
VN1101WBSD01


Tgt Ion: 56 Resp: 34770
Ion Ratio Lower Upper
56 100
69 31.5 25.2 37.8
84 92.2 77.8 116.8



#32
1,1,1-Trichloroethane
Concen: 19.681 ug/l
RT: 8.177 min Scan# 1041
Delta R.T. 0.006 min
Lab File: VN075102.D
Acq: 01 Nov 2022 12:33

Tgt Ion: 97 Resp: 36928
Ion Ratio Lower Upper
97 100
99 65.8 50.2 75.2
61 48.3 37.0 55.6





#33

1,2-Dichloroethane-d4

Concen: 55.482 ug/l

RT: 8.582 min Scan# 1111

Delta R.T. -0.006 min

Lab File: VN075102.D

Acq: 01 Nov 2022 12:33

Instrument :

MSVOA_N

ClientSampleId :





VN1101WBSD01

Tgt Ion: 65 Resp: 65837

Ion Ratio Lower Upper

65 100

67 50.6 0.0 102.0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.106 min Scan# 1199

Delta R.T. -0.000 min

Lab File: VN075102.D

Acq: 01 Nov 2022 12:33




Tgt Ion: 114 Resp: 165188


Ion Ratio Lower Upper

114 100

63 20.8 0.0 39.6

88 16.1 0.0 29.0









#35
Dibromofluoromethane
Concen: 52.183 ug/l
RT: 8.177 min Scan# 1040
Delta R.T. 0.006 min
Lab File: VN075102.D
Acq: 01 Nov 2022 12:33




Instrument :
MSVOA_N
ClientSampleId :
VN1101WBSD01


Tgt Ion:113 Resp: 50604
Ion Ratio Lower Upper
113 100
111 104.1 80.5 120.7
192 16.3 13.2 19.8



#36
1,1-Dichloropropene
Concen: 19.111 ug/l
RT: 8.377 min Scan# 1075
Delta R.T. -0.000 min
Lab File: VN075102.D
Acq: 01 Nov 2022 12:33

Tgt Ion: 75 Resp: 28692
Ion Ratio Lower Upper
75 100
110 35.4 17.8 53.3
77 31.6 24.6 37.0









#37
Ethyl Acetate
Concen: 20.091 ug/l
RT: 7.565 min Scan# 91
Delta R.T. -0.006 min
Lab File: VN075102.D
Acq: 01 Nov 2022 12:33




Instrument :
MSVOA_N
ClientSampleId :
VN1101WBSD01


Tgt Ion: 43 Resp: 33080
Ion Ratio Lower Upper
43 100
61 14.7 12.2 18.4
70 10.4 9.4 14.2



#38
Carbon Tetrachloride
Concen: 18.952 ug/l
RT: 8.371 min Scan# 1074
Delta R.T. -0.000 min
Lab File: VN075102.D
Acq: 01 Nov 2022 12:33

Tgt Ion: 117 Resp: 31299
Ion Ratio Lower Upper
117 100
119 95.5 76.7 115.1
121 32.3 23.9 35.9









#39
Methylcyclohexane
Concen: 18.742 ug/l
RT: 9.600 min Scan# 111
Delta R.T. -0.006 min
Lab File: VN075102.D
Acq: 01 Nov 2022 12:33




Instrument :
MSVOA_N
ClientSampleId :
VN1101WBSD01


Tgt Ion: 83 Resp: 37147
Ion Ratio Lower Upper
83 100
55 65.3 53.6 80.4
98 46.8 41.1 61.7



#40
Benzene
Concen: 19.596 ug/l
RT: 8.612 min Scan# 1115
Delta R.T. -0.000 min
Lab File: VN075102.D
Acq: 01 Nov 2022 12:33

Tgt Ion: 78 Resp: 90081
Ion Ratio Lower Upper
78 100
77 22.5 19.4 29.0









#41
Methacrylonitrile
Concen: 21.838 ug/l
RT: 7.794 min Scan# 91
Delta R.T. 0.006 min
Lab File: VN075102.D
Acq: 01 Nov 2022 12:33




Instrument :
MSVOA_N
ClientSampleId :
VN1101WBSD01


Tgt Ion:	41	Resp:	17226
Ion	Ratio	Lower	Upper
41	100		
39	58.7	44.4	66.6
67	82.4	69.8	104.6
52	35.6	27.3	40.9



#42
1,2-Dichloroethane
Concen: 20.617 ug/l
RT: 8.677 min Scan# 1126
Delta R.T. -0.000 min
Lab File: VN075102.D
Acq: 01 Nov 2022 12:33

Tgt Ion:	62	Resp:	34480
Ion	Ratio	Lower	Upper
62	100		
98	9.6	0.0	18.2









#43
Isopropyl Acetate
Concen: 20.935 ug/l
RT: 8.694 min Scan# 1129
Delta R.T. 0.000 min
Lab File: VN075102.D
Acq: 01 Nov 2022 12:33




Instrument :
MSVOA_N
ClientSampleId :
VN1101WBSD01


Tgt Ion: 43 Resp: 53467
Ion Ratio Lower Upper
43 100
61 26.7 22.4 33.6
87 15.9 12.5 18.7



#44
Trichloroethene
Concen: 19.284 ug/l
RT: 9.359 min Scan# 1242
Delta R.T. 0.006 min
Lab File: VN075102.D
Acq: 01 Nov 2022 12:33

Tgt Ion:130 Resp: 22240
Ion Ratio Lower Upper
130 100
95 101.9 0.0 198.4





#45

1,2-Dichloropropane

Concen: 21.308 ug/l

RT: 9.624 min Scan# 11

Delta R.T. -0.000 min

Lab File: VN075102.D

Acq: 01 Nov 2022 12:33

Instrument :

MSVOA_N

ClientSampleId :





VN1101WBSD01

Tgt Ion: 63 Resp: 23310

Ion Ratio Lower Upper

63 100

65 29.9 26.1 39.1



#46

Dibromomethane

Concen: 20.485 ug/l

RT: 9.712 min Scan# 1302

Delta R.T. -0.000 min

Lab File: VN075102.D

Acq: 01 Nov 2022 12:33




Tgt Ion: 93 Resp: 17459


Ion Ratio Lower Upper

93 100

95 85.9 66.6 99.8

174 84.2 70.3 105.5









#47
Bromodichloromethane
Concen: 20.180 ug/l
RT: 9.894 min Scan# 1333
Delta R.T. 0.000 min
Lab File: VN075102.D
Acq: 01 Nov 2022 12:33




Instrument :
MSVOA_N
ClientSampleId :
VN1101WBSD01


Tgt Ion: 83 Resp: 33307
Ion Ratio Lower Upper
83 100
85 68.3 52.1 78.1
127 9.2 6.6 10.0



#48
Methyl methacrylate
Concen: 20.999 ug/l
RT: 9.682 min Scan# 1297
Delta R.T. 0.000 min
Lab File: VN075102.D
Acq: 01 Nov 2022 12:33

Tgt Ion: 41 Resp: 24601
Ion Ratio Lower Upper
41 100
69 97.2 80.2 120.2
39 68.2 56.2 84.2





#49

1,4-Dioxane

Concen: 415.616 ug/l

RT: 9.694 min Scan# 1299

Delta R.T. -0.006 min

Lab File: VN075102.D

Acq: 01 Nov 2022 12:33

Instrument :

MSVOA_N

ClientSampleId :

VN1101WBSD01





Tgt Ion: 88 Resp: 12750

Ion Ratio Lower Upper

88 100

43 23.8 17.5 26.3

58 64.2 49.8 74.8



#50

Toluene-d8

Concen: 51.592 ug/l

RT: 10.571 min Scan# 1448

Delta R.T. -0.000 min

Lab File: VN075102.D




Acq: 01 Nov 2022 12:33


Tgt Ion: 98 Resp: 194105

Ion Ratio Lower Upper

98 100

100 65.7 53.0 79.6





#51

4-Methyl-2-Pentanone

Concen: 105.394 ug/l

RT: 10.447 min Scan# 1427

Delta R.T. 0.000 min

Lab File: VN075102.D

Acq: 01 Nov 2022 12:33

Instrument :

MSVOA_N

ClientSampleId :





VN1101WBSD01

Tgt Ion: 43 Resp: 173694

Ion Ratio Lower Upper

43 100

58 41.5 33.6 50.4



#52

Toluene

Concen: 19.491 ug/l

RT: 10.635 min Scan# 1459

Delta R.T. 0.000 min

Lab File: VN075102.D




Acq: 01 Nov 2022 12:33


Tgt Ion: 92 Resp: 59892

Ion Ratio Lower Upper

92 100

91 174.3 135.4 203.2









#53
 t-1,3-Dichloropropene
 Concen: 19.651 ug/l
 RT: 10.841 min Scan# 1494
 Delta R.T. 0.000 min
 Lab File: VN075102.D
 Acq: 01 Nov 2022 12:33




Instrument :
 MSVOA_N
 ClientSampleId :
 VN1101WBSD01


Tgt Ion: 75 Resp: 34234
 Ion Ratio Lower Upper
 75 100
 77 28.6 25.4 38.2



#54
 cis-1,3-Dichloropropene
 Concen: 19.986 ug/l
 RT: 10.312 min Scan# 1404
 Delta R.T. -0.006 min
 Lab File: VN075102.D
 Acq: 01 Nov 2022 12:33

Tgt Ion: 75 Resp: 36929
 Ion Ratio Lower Upper
 75 100
 77 28.3 23.7 35.5
 39 45.2 35.8 53.6





#55

1,1,2-Trichloroethane

Concen: 20.081 ug/l

RT: 11.024 min Scan# 11

Delta R.T. 0.005 min

Lab File: VN075102.D

Acq: 01 Nov 2022 12:33

Instrument :

MSVOA_N

ClientSampleId :

VN1101WBSD01

Tgt Ion: 97 Resp: 24467





Ion Ratio Lower Upper

97 100

83 88.1 65.9 98.9

85 52.8 40.4 60.6

99 68.1 46.4 69.6



#56

Ethyl methacrylate

Concen: 19.724 ug/l

RT: 10.876 min Scan# 1500

Delta R.T. -0.001 min

Lab File: VN075102.D

Acq: 01 Nov 2022 12:33




Tgt Ion: 69 Resp: 36658


Ion Ratio Lower Upper

69 100

41 59.7 45.6 68.4

39 39.8 28.5 42.7









#57
1,3-Dichloropropane
Concen: 20.856 ug/l
RT: 11.165 min Scan# 11
Delta R.T. -0.000 min
Lab File: VN075102.D
Acq: 01 Nov 2022 12:33




Instrument :
MSVOA_N
ClientSampleId :
VN1101WBSD01


Tgt Ion: 76 Resp: 42318
Ion Ratio Lower Upper
76 100
78 32.5 25.8 38.6



#58
2-Chloroethyl Vinyl ether
Concen: 108.169 ug/l
RT: 10.165 min Scan# 1379
Delta R.T. -0.000 min
Lab File: VN075102.D
Acq: 01 Nov 2022 12:33

Tgt Ion: 63 Resp: 77720
Ion Ratio Lower Upper
63 100
106 29.8 24.8 37.2









#59
2-Hexanone
Concen: 104.741 ug/l
RT: 11.200 min Scan# 1555
Delta R.T. -0.000 min
Lab File: VN075102.D
Acq: 01 Nov 2022 12:33




Instrument :
MSVOA_N
ClientSampleId :
VN1101WBSD01


Tgt Ion: 43 Resp: 131010
Ion Ratio Lower Upper
43 100
58 58.0 29.6 88.8



#60
Dibromochloromethane
Concen: 19.866 ug/l
RT: 11.359 min Scan# 1582
Delta R.T. -0.000 min
Lab File: VN075102.D
Acq: 01 Nov 2022 12:33

Tgt Ion:129 Resp: 25537
Ion Ratio Lower Upper
129 100
127 77.6 38.6 115.8









#61
1,2-Dibromoethane
Concen: 20.064 ug/l
RT: 11.471 min Scan# 1601
Delta R.T. -0.000 min
Lab File: VN075102.D
Acq: 01 Nov 2022 12:33




Instrument :
MSVOA_N
ClientSampleId :
VN1101WBSD01


Tgt Ion:107 Resp: 25053
Ion Ratio Lower Upper
107 100
109 95.8 75.6 113.4



#62
4-Bromofluorobenzene
Concen: 51.163 ug/l
RT: 12.853 min Scan# 1836
Delta R.T. -0.000 min
Lab File: VN075102.D
Acq: 01 Nov 2022 12:33

Tgt Ion: 95 Resp: 67619
Ion Ratio Lower Upper
95 100
174 74.1 0.0 147.4
176 72.2 0.0 142.2









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 11
Delta R.T. -0.006 min
Lab File: VN075102.D
Acq: 01 Nov 2022 12:33




Instrument :
MSVOA_N
ClientSampleId :
VN1101WBSD01


Tgt Ion:117 Resp: 150176
Ion Ratio Lower Upper
117 100
82 55.3 43.1 64.7
119 31.8 25.5 38.3



#64
Tetrachloroethene
Concen: 21.232 ug/l
RT: 11.106 min Scan# 1539
Delta R.T. -0.000 min
Lab File: VN075102.D
Acq: 01 Nov 2022 12:33

Tgt Ion:164 Resp: 19548
Ion Ratio Lower Upper
164 100
166 123.0 106.5 159.7
129 101.3 81.9 122.9
131 94.0 81.4 122.2





#65

Chlorobenzene

Concen: 20.228 ug/l

RT: 11.894 min Scan# 1673

Delta R.T. 0.000 min

Lab File: VN075102.D

Acq: 01 Nov 2022 12:33

Instrument :

MSVOA_N

ClientSampleId :





VN1101WBSD01

Tgt Ion:112 Resp: 64841

Ion Ratio Lower Upper

112 100

114 30.9 25.4 38.2



#66

1,1,1,2-Tetrachloroethane

Concen: 19.820 ug/l

RT: 11.965 min Scan# 1685

Delta R.T. -0.000 min

Lab File: VN075102.D

Acq: 01 Nov 2022 12:33




Tgt Ion:131 Resp: 23568


Ion Ratio Lower Upper

131 100

133 94.7 47.8 143.4

119 65.9 33.0 99.0









#67
Ethyl Benzene
Concen: 19.687 ug/l
RT: 11.965 min Scan# 11
Delta R.T. -0.000 min
Lab File: VN075102.D
Acq: 01 Nov 2022 12:33




Instrument :
MSVOA_N
ClientSampleId :
VN1101WBSD01


Tgt Ion: 91 Resp: 113537
Ion Ratio Lower Upper
91 100
106 30.5 24.9 37.3



#68
m/p-Xylenes
Concen: 38.857 ug/l
RT: 12.070 min Scan# 1703
Delta R.T. -0.006 min
Lab File: VN075102.D
Acq: 01 Nov 2022 12:33

Tgt Ion:106 Resp: 89918
Ion Ratio Lower Upper
106 100
91 203.3 156.0 234.0









#69
o-Xylene
Concen: 19.543 ug/l
RT: 12.400 min Scan# 1759
Delta R.T. -0.000 min
Lab File: VN075102.D
Acq: 01 Nov 2022 12:33




Instrument :
MSVOA_N
ClientSampleId :
VN1101WBSD01


Tgt Ion:106 Resp: 44725
Ion Ratio Lower Upper
106 100
91 213.9 105.2 315.5



#70
Styrene
Concen: 19.513 ug/l
RT: 12.412 min Scan# 1761
Delta R.T. -0.000 min
Lab File: VN075102.D
Acq: 01 Nov 2022 12:33

Tgt Ion:104 Resp: 71204
Ion Ratio Lower Upper
104 100
78 49.4 38.1 57.1
103 54.2 44.6 66.8





#71

Bromoform

Concen: 19.203 ug/l

RT: 12.582 min Scan# 1

Delta R.T. 0.000 min

Lab File: VN075102.D

Acq: 01 Nov 2022 12:33

Instrument :

MSVOA_N

ClientSampleId :

VN1101WBSD01





Tgt Ion:173 Resp: 17096

Ion Ratio Lower Upper

173 100

175 47.3 24.6 74.0

254 0.0 0.0 0.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.794 min Scan# 1996

Delta R.T. -0.000 min

Lab File: VN075102.D

Acq: 01 Nov 2022 12:33




Tgt Ion:152 Resp: 72181


Ion Ratio Lower Upper

152 100

115 62.2 30.2 90.6

150 166.9 0.0 348.4









#73
Isopropylbenzene
Concen: 19.451 ug/l
RT: 12.700 min Scan# 1810
Delta R.T. -0.000 min
Lab File: VN075102.D
Acq: 01 Nov 2022 12:33




Instrument :
MSVOA_N
ClientSampleId :
VN1101WBSD01


Tgt Ion:105 Resp: 112716
Ion Ratio Lower Upper
105 100
120 26.1 13.1 39.1



#74
N-ethyl acetate
Concen: 20.705 ug/l
RT: 12.494 min Scan# 1775
Delta R.T. 0.000 min
Lab File: VN075102.D
Acq: 01 Nov 2022 12:33

Tgt Ion: 43 Resp: 44376
Ion Ratio Lower Upper
43 100
70 47.9 39.8 59.8
55 26.7 19.9 29.9
61 25.6 21.7 32.5





#75

1,1,2,2-Tetrachloroethane

Concen: 21.887 ug/l

RT: 12.935 min Scan# 1850

Delta R.T. -0.006 min

Lab File: VN075102.D

Acq: 01 Nov 2022 12:33

Instrument :

MSVOA_N

ClientSampleId :

VN1101WBSD01





Tgt Ion: 83 Resp: 40181

Ion Ratio Lower Upper

83 100

131 9.6 5.1 15.3

85 63.5 31.6 94.7



#76

1,2,3-Trichloropropane

Concen: 31.930 ug/l

RT: 12.994 min Scan# 1860

Delta R.T. -0.000 min

Lab File: VN075102.D




Acq: 01 Nov 2022 12:33


Tgt Ion: 75 Resp: 48936

Ion Ratio Lower Upper

75 100

77 116.0 75.6 226.8





#77

Bromobenzene

Concen: 19.856 ug/l

RT: 12.982 min Scan# 1858

Delta R.T. 0.000 min

Lab File: VN075102.D

Acq: 01 Nov 2022 12:33

Instrument :

MSVOA_N

ClientSampleId :

VN1101WBSD01





Tgt Ion:156 Resp: 25335

Ion Ratio Lower Upper

156 100

77 224.1 108.5 325.5

158 98.8 47.7 143.1



#78

n-propylbenzene

Concen: 19.894 ug/l

RT: 13.035 min Scan# 1867

Delta R.T. 0.000 min

Lab File: VN075102.D




Acq: 01 Nov 2022 12:33


Tgt Ion: 91 Resp: 129431

Ion Ratio Lower Upper

91 100

120 22.5 11.9 35.5









#79
2-Chlorotoluene
Concen: 20.292 ug/l
RT: 13.129 min Scan# 1883
Delta R.T. 0.000 min
Lab File: VN075102.D
Acq: 01 Nov 2022 12:33




Instrument :
MSVOA_N
ClientSampleId :
VN1101WBSD01


Tgt Ion: 91 Resp: 80674
Ion Ratio Lower Upper
91 100
126 34.3 16.9 50.7



#80
1,3,5-Trimethylbenzene
Concen: 20.188 ug/l
RT: 13.176 min Scan# 1891
Delta R.T. 0.000 min
Lab File: VN075102.D
Acq: 01 Nov 2022 12:33

Tgt Ion: 105 Resp: 98257
Ion Ratio Lower Upper
105 100
120 48.0 24.7 74.1









#81
trans-1,4-Dichloro-2-butene
Concen: 18.123 ug/l
RT: 12.741 min Scan# 1817
Delta R.T. 0.000 min
Lab File: VN075102.D
Acq: 01 Nov 2022 12:33




Instrument :
MSVOA_N
ClientSampleId :
VN1101WBSD01


Tgt Ion: 75 Resp: 10820
Ion Ratio Lower Upper
75 100
53 103.6 78.5 117.7
89 51.6 37.7 56.5



#82
4-Chlorotoluene
Concen: 20.292 ug/l
RT: 13.129 min Scan# 1883
Delta R.T. 0.000 min
Lab File: VN075102.D
Acq: 01 Nov 2022 12:33

Tgt Ion: 91 Resp: 80674
Ion Ratio Lower Upper
91 100
126 34.3 16.9 50.7









#83
 tert-Butylbenzene
 Concen: 19.697 ug/l
 RT: 13.441 min Scan# 1936
 Delta R.T. 0.000 min
 Lab File: VN075102.D
 Acq: 01 Nov 2022 12:33




Instrument :
 MSVOA_N
 ClientSampleId :
 VN1101WBSD01


Tgt Ion:119 Resp: 85187
 Ion Ratio Lower Upper
 119 100
 91 67.5 32.8 98.4
 134 27.1 12.3 36.9



#84
 1,2,4-Trimethylbenzene
 Concen: 19.975 ug/l
 RT: 13.482 min Scan# 1943
 Delta R.T. 0.000 min
 Lab File: VN075102.D
 Acq: 01 Nov 2022 12:33

Tgt Ion:105 Resp: 98368
 Ion Ratio Lower Upper
 105 100
 120 47.0 22.7 68.1









#85
 sec-Butylbenzene
 Concen: 19.667 ug/l
 RT: 13.617 min Scan# 1966
 Delta R.T. -0.001 min
 Lab File: VN075102.D
 Acq: 01 Nov 2022 12:33




Instrument :
 MSVOA_N
 ClientSampleId :
 VN1101WBSD01


Tgt Ion:105 Resp: 118985
 Ion Ratio Lower Upper
 105 100
 134 19.1 10.2 30.6



#86
 p-Isopropyltoluene
 Concen: 19.589 ug/l
 RT: 13.729 min Scan# 1985
 Delta R.T. -0.006 min
 Lab File: VN075102.D
 Acq: 01 Nov 2022 12:33

Tgt Ion:119 Resp: 97878
 Ion Ratio Lower Upper
 119 100
 134 26.1 13.1 39.3
 91 24.2 11.9 35.5









#87
 1,3-Dichlorobenzene
 Concen: 19.541 ug/l
 RT: 13.735 min Scan# 1986
 Delta R.T. 0.000 min
 Lab File: VN075102.D
 Acq: 01 Nov 2022 12:33




Instrument :
 MSVOA_N
 ClientSampleId :
 VN1101WBSD01


Tgt Ion:146 Resp: 48819
 Ion Ratio Lower Upper
 146 100
 111 42.5 20.8 62.5
 148 65.0 31.1 93.5



#88
 1,4-Dichlorobenzene
 Concen: 20.084 ug/l
 RT: 13.812 min Scan# 1999
 Delta R.T. -0.000 min
 Lab File: VN075102.D
 Acq: 01 Nov 2022 12:33

Tgt Ion:146 Resp: 49656
 Ion Ratio Lower Upper
 146 100
 111 40.8 20.1 60.3
 148 63.5 31.9 95.7









#89
n-Butylbenzene
Concen: 19.906 ug/l
RT: 14.059 min Scan# 2041
Delta R.T. -0.000 min
Lab File: VN075102.D
Acq: 01 Nov 2022 12:33




Instrument :
MSVOA_N
ClientSampleId :
VN1101WBSD01


Tgt Ion: 91 Resp: 80095
Ion Ratio Lower Upper
91 100
92 52.7 26.5 79.4
134 24.9 13.0 39.0



#90
Hexachloroethane
Concen: 18.777 ug/l
RT: 14.335 min Scan# 2088
Delta R.T. 0.000 min
Lab File: VN075102.D
Acq: 01 Nov 2022 12:33

Tgt Ion: 117 Resp: 17202
Ion Ratio Lower Upper
117 100
201 59.7 31.5 94.5









#91
1,2-Dichlorobenzene
Concen: 20.160 ug/l
RT: 14.106 min Scan# 2049
Delta R.T. -0.000 min
Lab File: VN075102.D
Acq: 01 Nov 2022 12:33




Instrument :
MSVOA_N
ClientSampleId :
VN1101WBSD01


Tgt Ion:146 Resp: 50069
Ion Ratio Lower Upper
146 100
111 44.1 21.3 63.9
148 64.8 31.4 94.2



#92
1,2-Dibromo-3-Chloropropane
Concen: 21.048 ug/l
RT: 14.723 min Scan# 2154
Delta R.T. 0.006 min
Lab File: VN075102.D
Acq: 01 Nov 2022 12:33

Tgt Ion: 75 Resp: 8227
Ion Ratio Lower Upper
75 100
155 67.6 42.1 126.3
157 93.6 51.7 155.1





#93

1,2,4-Trichlorobenzene

Concen: 19.105 ug/l

RT: 15.394 min Scan# 2268

Delta R.T. -0.000 min

Lab File: VN075102.D

Acq: 01 Nov 2022 12:33

Instrument :

MSVOA_N

ClientSampleId :

VN1101WBSD01





Tgt Ion:180 Resp: 20418

Ion Ratio Lower Upper

180 100

182 97.7 47.8 143.4

145 35.0 16.8 50.3



#94

Hexachlorobutadiene

Concen: 19.220 ug/l

RT: 15.506 min Scan# 2287

Delta R.T. 0.006 min

Lab File: VN075102.D

Acq: 01 Nov 2022 12:33




Tgt Ion:225 Resp: 10014


Ion Ratio Lower Upper

225 100

223 62.6 32.3 96.9

227 62.4 32.0 96.2









#95
Naphthalene
Concen: 19.294 ug/l
RT: 15.641 min Scan# 2310
Delta R.T. -0.000 min
Lab File: VN075102.D
Acq: 01 Nov 2022 12:33

Instrument :
MSVOA_N
ClientSampleId :
VN1101WBSD01

Tgt Ion:	128	Resp:	79298
Ion Ratio	Lower	Upper	
128	100		
127	13.6	10.4	15.6
129	10.7	8.6	12.8



#96
1,2,3-Trichlorobenzene
Concen: 20.459 ug/l
RT: 15.841 min Scan# 2344
Delta R.T. -0.000 min
Lab File: VN075102.D
Acq: 01 Nov 2022 12:33

Tgt Ion:	180	Resp:	21726
Ion Ratio	Lower	Upper	
180	100		
182	91.4	47.3	141.8
145	36.1	17.6	52.9

