

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110122\
 Data File : VN075104.D
 Acq On : 01 Nov 2022 17:58
 Operator : JC\MD
 Sample : N5417-09
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 AUD-1906

Quant Time: Nov 02 07:12:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	105674	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	198964	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	181029	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	75197	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	81631	57.332	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	114.660%
35) Dibromofluoromethane	8.171	113	64601	55.308	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	110.620%
50) Toluene-d8	10.571	98	248325	54.799	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	109.600%
62) 4-Bromofluorobenzene	12.847	95	78987	49.619	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.240%
Target Compounds						
					Qvalue	
16) Acetone	4.448	43	8520	14.132	ug/l	100
25) 2-Butanone	7.494	43	37120	39.145	ug/l	97
52) Toluene	10.635	92	258626	69.877	ug/l	96
80) 1,3,5-Trimethylbenzene	13.176	105	5518	1.088	ug/l	98
84) 1,2,4-Trimethylbenzene	13.488	105	14676	2.861	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110122\
Data File : VN075104.D
Acq On : 01 Nov 2022 17:58
Operator : JC\MD
Sample : N5417-09
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
AUD-1906

Quant Time: Nov 02 07:12:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

