

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110123\
 Data File : VN079775.D
 Acq On : 01 Nov 2023 16:44
 Operator : JC\MD
 Sample : 05173-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-188-GW-150-152

Quant Time: Nov 02 05:56:42 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 02 05:49:08 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.988	168	333	50.000	ug/l	#-0.24
34) 1,4-Difluorobenzene	8.865	114	662366	50.000	ug/l	-0.24
63) Chlorobenzene-d5	11.818	117	362441	50.000	ug/l	-0.05
72) 1,4-Dichlorobenzene-d4	13.782	152	176190	50.000	ug/l	-0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.265	65	280826	55352.985	ug/l	-0.32
Spiked Amount	50.000	Range 74 - 125	Recovery	= 110705.960%#		
35) Dibromofluoromethane	7.800	113	208722	44.523	ug/l	-0.37
Spiked Amount	50.000	Range 75 - 124	Recovery	= 89.040%		
50) Toluene-d8	10.453	98	779777	45.183	ug/l	-0.12
Spiked Amount	50.000	Range 86 - 113	Recovery	= 90.360%		
62) 4-Bromofluorobenzene	0.000	95	0	0.000	ug/l	
Spiked Amount	50.000	Range 83 - 123	Recovery	= 0.000%#		
Target Compounds						Qvalue
38) Carbon Tetrachloride	7.882	117	30914	4.583	ug/l #	17
44) Trichloroethene	9.141	130	1867	0.375	ug/l	55
55) 1,1,2-Trichloroethane	10.465	97	19229	4.022	ug/l #	1
56) Ethyl methacrylate	10.459	69	2339	0.343	ug/l #	45
60) Dibromochloromethane	11.023	129	1529	0.299	ug/l #	11
64) Tetrachloroethene	11.018	164	1438	0.613	ug/l	90
71) Bromoform	12.829	173	1912	0.951	ug/l #	1
76) 1,2,3-Trichloropropane	12.823	75	101311	24.492	ug/l #	1
81) trans-1,4-Dichloro-2-b...	12.823	75	101311	66.638	ug/l #	15
95) Naphthalene	15.647	128	2524	0.272	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110123\
Data File : VN079775.D
Acq On : 01 Nov 2023 16:44
Operator : JC\MD
Sample : 05173-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-188-GW-150-152

Quant Time: Nov 02 05:56:42 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 02 05:49:08 2023
Response via : Initial Calibration

