

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN110218\
 Data File : VN052120.D
 Acq On : 2 Nov 2018 11:14
 Operator : MD\SY
 Sample : PB114370ZHE#09
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 7 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB114370ZHE#09

Manual Integrations
 APPROVED

MMDadoda
 11/5/2018 11:52:56 AM

Quant Time: Nov 02 22:49:45 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N103018W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 31 02:45:22 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	298668	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	558456	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	526016	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	216267	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	252195	52.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.24%	
35) Dibromofluoromethane	7.59	113	209057	59.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	119.68%	
50) Toluene-d8	10.09	98	833383	59.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	119.36%	
62) 4-Bromofluorobenzene	12.40	95	281943	55.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.00%	
Target Compounds						
16) Acetone	3.82	43	13142	12.846	ug/l #	87
20) Methylene Chloride	4.57	84	7133m	1.830	ug/l	
43) Isopropyl Acetate	8.17	43	194194	28.727	ug/l	99
95) Naphthalene	15.13	128	6750	1.089	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN110218\
Data File : VN052120.D
Acq On : 2 Nov 2018 11:14
Operator : MD\SY
Sample : PB114370ZHE#09
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 7 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB114370ZHE#09

Manual Integrations
APPROVED

MMDadoda
11/5/2018 11:52:56 AM

Quant Time: Nov 02 22:49:45 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N103018W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 31 02:45:22 2018
Response via : Initial Calibration

