

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN110218\
 Data File : VN052121.D
 Acq On : 2 Nov 2018 11:41
 Operator : MD\SY
 Sample : PB114370ZHE#10
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 8 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB114370ZHE#10

Manual Integrations
 APPROVED

MMDadoda
 11/5/2018 11:53:01 AM

Quant Time: Nov 02 22:52:15 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N103018W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 31 02:45:22 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	303181	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	561738	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	541253	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	224303	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	258625	53.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.32%	
35) Dibromofluoromethane	7.59	113	219783	62.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	125.08%	
50) Toluene-d8	10.09	98	848239	60.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	120.78%	
62) 4-Bromofluorobenzene	12.40	95	289068	56.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.12%	
Target Compounds						
16) Acetone	3.83	43	13976m	13.458	ug/l	99
20) Methylene Chloride	4.57	84	6019m	1.521	ug/l	
43) Isopropyl Acetate	8.17	43	190888	28.073	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN110218\
Data File : VN052121.D
Acq On : 2 Nov 2018 11:41
Operator : MD\SY
Sample : PB114370ZHE#10
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 8 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB114370ZHE#10

Manual Integrations
APPROVED

MMDadoda
11/5/2018 11:53:01 AM

Quant Time: Nov 02 22:52:15 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N103018W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 31 02:45:22 2018
Response via : Initial Calibration

