

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN110218\
 Data File : VN052122.D
 Acq On : 2 Nov 2018 12:07
 Operator : MD\SY
 Sample : PB114370ZHE#11
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB114370ZHE#11

Quant Time: Nov 02 22:54:33 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N103018W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 31 02:45:22 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	319085	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	593932	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	569120	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	244244	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	271972	53.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.24%	
35) Dibromofluoromethane	7.59	113	223181	60.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	120.14%	
50) Toluene-d8	10.09	98	891067	60.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	120.00%	
62) 4-Bromofluorobenzene	12.40	95	312711	57.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.72%	
Target Compounds						
16) Acetone	3.82	43	14271	13.057	ug/l	97
20) Methylene Chloride	4.56	84	7410	1.779	ug/l #	74
43) Isopropyl Acetate	8.17	43	193408	26.902	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN110218\
Data File : VN052122.D
Acq On : 2 Nov 2018 12:07
Operator : MD\SY
Sample : PB114370ZHE#11
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 9 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB114370ZHE#11

Quant Time: Nov 02 22:54:33 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N103018W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 31 02:45:22 2018
Response via : Initial Calibration

