

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN110218\
 Data File : VN052126.D
 Acq On : 2 Nov 2018 13:55
 Operator : MD\SY
 Sample : PB114370ZHE#15
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB114370ZHE#15

Quant Time: Nov 02 23:15:00 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N103018W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 31 02:45:22 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	279598	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	532427	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	499258	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	207424	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	242423	54.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.06%	
35) Dibromofluoromethane	7.59	113	197110	59.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	118.36%	
50) Toluene-d8	10.09	98	798061	59.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	119.90%	
62) 4-Bromofluorobenzene	12.40	95	270068	55.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.52%	
Target Compounds						
16) Acetone	3.82	43	14699	15.348	ug/l #	86
20) Methylene Chloride	4.56	84	6946	1.903	ug/l #	77
43) Isopropyl Acetate	8.17	43	191218	29.670	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN110218\
Data File : VN052126.D
Acq On : 2 Nov 2018 13:55
Operator : MD\SY
Sample : PB114370ZHE#15
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB114370ZHE#15

Quant Time: Nov 02 23:15:00 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N103018W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 31 02:45:22 2018
Response via : Initial Calibration

