

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN110218\
 Data File : VN052133.D
 Acq On : 2 Nov 2018 17:04
 Operator : MD\SY
 Sample : J5741-50
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 20 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-12

Quant Time: Nov 02 23:41:16 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N103018W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 31 02:45:22 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	273145	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	506931	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	484940	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	189173	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	238353	54.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.76%	
35) Dibromofluoromethane	7.60	113	192464	60.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	121.38%	
50) Toluene-d8	10.09	98	772025	60.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	121.82%	
62) 4-Bromofluorobenzene	12.40	95	258604	55.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.16%	
Target Compounds						
16) Acetone	3.82	43	9220	9.855	ug/l	94
20) Methylene Chloride	4.56	84	909175	255.026	ug/l	97
43) Isopropyl Acetate	8.17	43	182531	29.746	ug/l	97
95) Naphthalene	15.14	128	9270	1.710	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN110218\
Data File : VN052133.D
Acq On : 2 Nov 2018 17:04
Operator : MD\SY
Sample : J5741-50
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 20 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
TP-12

Quant Time: Nov 02 23:41:16 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N103018W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 31 02:45:22 2018
Response via : Initial Calibration

