


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN110218\
Data File : VN052114.D
Acq On : 2 Nov 2018 8:16
Operator : MD\SY
Sample : BFB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 1 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N103018W.M
Title : SW846 8260
Last Update : Wed Oct 31 02:45:22 2018



AutoFind: Scans 3305, 3306, 3307; Background Corrected with Scan 3291

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.7	15604	PASS
75	95	30	60	50.1	37781	PASS
95	95	100	100	100.0	75442	PASS
96	95	5	9	6.4	4851	PASS
173	174	0.00	2	1.6	854	PASS
174	95	50	100	68.8	51936	PASS
175	174	5	9	8.3	4327	PASS
176	174	95	101	96.9	50309	PASS
177	176	5	9	6.9	3468	PASS