

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN110220\
 Data File : VN064418.D
 Acq On : 2 Nov 2020 16:42
 Operator : JC/MD
 Sample : VN1102WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1102WBL01

Quant Time: Nov 02 17:09:01 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N110220W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Mon Nov 02 15:45:48 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.15	128	39411	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.55	114	215908	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.38	117	184011	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	7.98	65	99928	31.06	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	103.53%
60) 4-Bromofluorobenzene	12.38	95	79994	25.72	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	85.73%
63) Toluene-d8	10.06	98	251737	30.68	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	102.27%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN110220\
Data File : VN064418.D
Acq On : 2 Nov 2020 16:42
Operator : JC/MD
Sample : VN1102WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1102WBL01

Quant Time: Nov 02 17:09:01 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA N\METHODS\624N110220W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Mon Nov 02 15:45:48 2020
Response via : Initial Calibration

